



Since 1998

NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
E-mail : ncqc@calibrationlaboratory.in ; calibrationlab.ncqc@gmail.com

Visit our Web Site : www.calibrationlaboratory.in



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/T/O 0/03 Issue No. 02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No. NCQC-T/220223/005	
	Date of Issue 23/02/2023	
	Date of Calibration 22/02/2023	
	Suggested Due Date 21/02/2024	
Date of Receipt / Ref.No.: 16/02/2023		

ULR No. : CC21282300000799F	Discipline : Thermal Calibration Specific Heat and Humidity
------------------------------------	---

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T – 155
	Serial No. : 3354
	Name of Instrument : Temperature & Humidity Data Logger

Range 0 to 99 % Rh	Visual Inspection Accuracy OK $\pm 5\% \text{ Rh}$
Resolution 1 % Rh	
Make / Model Elitech/RC-4HC	
Location Laboratory	

Loc. of Calibration At Laboratory	Std. Ref. Used Cali.Method No. NCQC/CM/T/010
Environment Cond. 25 \pm 4 $^{\circ}\text{C}$, 30 to 75 % Rh	

Master Instruments Detail

NCQC SYS.NO.	Nomenclature	Make/ Model	Id / Serial No.	Traceable To	Certificate No.	Valid Upto
335	Digital thermo Hygro Palm	Rotronic / HP32	NCQC/T-161/ 5181935 / 20281776	CC 2733	30056028	05/07/2023
435	4 wire RTD with Indicator	RTD - Tempensens / TPRT-100, Indicator - Tempensens / TEMPMET-09-02	RTD-NCQC/T-235 / 23000299; Indicator - NCQC/T-236 / T0064	CC-2840	TL/022/1068.2.1	25/09/2023

Calibration Results of Humidity @ 25 $^{\circ}\text{C}$

Set point on Hybrid Temperature and Humidity Calibrator in % Rh	Reading observed by UUC in % Rh	Reading observed by Master instrument in % Rh	Absolute Error in % Rh	Expanded Uncertainty in % Rh (k=2) \pm
20.0	20.6	20.05	0.55	3.20
40.0	40.9	40.22	0.68	3.20
50.0	51.2	50.25	0.95	3.20
70.0	71.5	70.33	1.17	3.20
95.0	96.8	95.46	1.34	3.20

NCQC
 Valid up to 21-02-2024
 Reviewed *Jigar Panchal*

Traceable To National / International Standards.	
Calibrated By <i>[Signature]</i> Chintan Patel Calibration Engineer	Reviewed & Approved By <i>[Signature]</i> Jigar Panchal Calibration Incharge

NCQC System Certificate No. 319



Since 1998

NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
E-mail : ncqc@calibrationlaboratory.in, calibrationlab.ncqc@gmail.com

Visit our Web Site : www.calibrationlaboratory.in



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/T/O 0/03 Issue No. 02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No.	NCQC-T/220223/005
	Date of Issue	23/02/2023
	Date of Calibration	22/02/2023
	Suggested Due Date	21/02/2024
	Date of Receipt / Ref.No.:	16/02/2023

ULR No. : CC212823000000799F	Discipline : Thermal Calibration Specific Heat and Humidity
------------------------------	--

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T – 155
	Serial No. : 3354
	Name of Instrument : Temperature & Humidity Data Logger

Range	-30 to 60 °C	Accuracy	± 0.5 °C up to 20 to 40°C & Other ±1°C
Least Count	0.1 °C		


Calibration Results of Temperature


Set point on Hybrid Temperature and Humidity Calibrator in °C	Reading observed by UUC in °C	Reading observed by Master instrument in °C	Absolute Error in °C	Expanded Uncertainty in °C (k = 2) ±
-25.0	-25.4	-25.224	0.176	0.78
-15.0	-15.3	-15.182	0.118	0.78
0.0	-0.1	-0.015	0.085	0.78
10.0	10.3	10.194	0.106	0.78
30.0	30.4	30.192	0.208	0.78
50.0	50.5	50.286	0.214	0.78

Remarks:

- Averages of minimum five readings are reported.
- Suggested due date is given based on customer requirements
- Calibration Points are given based on customer requirements.
- Condition of instrument found satisfactory during receipt.
- These results are obtained at the time of calibration.
- Any hand written corrections (except @) or photocopies of the report invalidates this certificate.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with National / international standard.
- Portable Hybrid temperature and Humidity calibrator was used as a source of generating temperature and humidity as well as controlling temperature at specified set point.

Traceable To National / International Standards.

Calibrated By  Chintan Patel
Calibration Engineer

Reviewed & Approved By  Jigar Panchal
Calibration Incharge