


TRUE COPY



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058

Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370

E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com

Visit our Web Site : www.calibrationlaboratory.in



Precision Calibration with National / International Traceability for Temperature, Dimension, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/T/10/03 Issue No. 02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No. NCQC-T/170223/001	Date of Issue 18/02/2023 Date of Calibration 17/02/2023 Suggested Due Date 16/02/2024
	Date of Receipt / Ref.No.: 16/02/2023	

ULR No. : CC212823000000710F	Discipline : Thermal Calibration Specific Heat and Humidity
-------------------------------------	---

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T – 162 Serial No. : EF7186H00537 Name of Instrument : Temperature & Humidity Data Logger
---	--

Range 0 to 99 % Rh Resolution 0.1 % Rh Make / Model Elitech/RC-4HC Location Laboratory	Visual Inspection Accuracy OK $\pm 5\% \text{ Rh}$
---	---

Loc. of Calibration Environment Cond. At Laboratory $25 \pm 4 \text{ }^\circ\text{C}$, 30 to 75 % Rh	Std. Ref. Used Cali.Method No. NCQC/CM/T/010
--	--

Master Instruments Detail

NCQC SYS.NO.	Nomenclature	Make/ Model	Id / Serial No.	Traceable To	Certificate No.	Valid Upto
335	Digital thermo Hygro Palm	Rotronic / HP32	NCQC/T-161/ 5181935 / 20281776	CC 2733	30056028	05/07/2023
435	4 wire RTD with Indicator	RTD - Tempsens / TPRT-100, Indicator - Tempsens / TEMPMET-09-02	RTD-NCQC/T-235 / 23000299, Indicator - NCQC/T-236 / T0064	CC-2840	TL/022/1068.2.1	25/09/2023

Calibration Results of Humidity @ 25 °C


Set point on Hybrid Temperature and Humidity Calibrator in % Rh	Reading observed by UUC in % Rh	Reading observed by Master instrument in % Rh	Absolute Error in % Rh	Expanded Uncertainty in % Rh (k=2) \pm
20.0	20.4	20.06	0.34	3.20
40.0	40.8	40.20	0.60	3.20
50.0	51.1	50.28	0.82	3.20
70.0	71.6	70.42	1.18	3.20
95.0	96.9	95.61	1.29	3.20

NCQC
 Valid up to 16-02-2024
 Reviewed: *Jigir*

Traceable To National / International Standards.

Calibrated By <i>Vishal Chauhan</i> Vishal Chauhan Calibration Engineer	Reviewed & Approved By <i>Jigir Panchal</i> Jigir Panchal Calibration Incharge
---	--

NCQC System Certificate No. 337



NATIONAL CENTRE FOR QUALITY CALIBRATION

TRUE COPY

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
 Ph. +91-79-29795322, 29795323 • Fax: +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
 E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com
 Visit our Web Site : www.calibrationlaboratory.in



Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/T/10/03 Issue No. 02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No. NCQC-T/170223/001	Date of Issue 18/02/2023 Date of Calibration 17/02/2023 Suggested Due Date 16/02/2024 Date of Receipt / Ref.No.: 16/02/2023
	Date of Issue 18/02/2023	
	Date of Calibration 17/02/2023	
	Suggested Due Date 16/02/2024	

ULR No. : CC212823000000710F **Discipline :** Thermal Calibration
 Specific Heat and Humidity

Details of Observation of Unit Under Calibration

Identification No. : NCQC/T – 162
Serial No. : EF7186H00537
Name of Instrument : Temperature & Humidity Data Logger

Range	-30 to 60 °C	Accuracy	± 0.5 °C up to - 20 to 40°C & Other ±1 °C
Least Count	0.1 °C		

Calibration Results of Temperature

Set point on Hybrid Temperature and Humidity Calibrator in °C	Reading observed by UUC in °C	Reading observed by Master instrument in °C	Absolute Error in °C	Expanded Uncertainty in °C (k =2) ±
-25.0	-25.4	-25.215	0.185	0.78
-15.0	-15.3	-15.122	0.178	0.78
0.0	0.0	0.025	0.025	0.78
10.0	10.2	10.077	0.123	0.78
30.0	30.3	30.165	0.135	0.78
50.0	50.4	50.208	0.192	0.78

Remarks:

- Averages of minimum five readings are reported.
- Suggested due date is given based on customer requirements
- Calibration Points are given based on customer requirements.
- Condition of instrument found satisfactory during receipt.
- These results are obtained at the time of calibration.
- Any hand written corrections (except @) or photocopies of the report invalidates this certificate.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with National / international standard.
- Portable Hybrid temperature and Humidity calibrator was used as a source of generating temperature and humidity as well as controlling temperature at specified set point.

Traceable To National / International Standards.

Calibrated By Vishal Chauhan Calibration Engineer	Reviewed & Approved By Jigar Panchal Calibration Incharge
--	--