



CALIBRATION CERTIFICATE

CUSTOMER : M/S NATIONAL CENTRE FOR QUALITY CALIBRATION
HOUSE NO. 4, ABHISHREE CORPORATE PARK, NR. SWAGAT BUNGALOW BRTS, ISKCON-
AMBLI ROAD, AMBLI, AHMEDABAD -380058

Date of Performance	: 09/09/2023	Certificate No.	: QSI/ 1523/23/09
Date of Issue	: 11/09/2023	Service Request Date / Receipt Date	: 08/09/2023
Service Request No.	: 1523	Total Page	: 1
Page	: 1	Environment Temperature	: (20 ± 1)°C
Unit Under Calibration	: OPTICAL PARALLEL SET	Relative Humidity(%RH)	: (50 ± 10) %
Calibration Procedure	: QMP-25(WI-80M)	Location of calibration	: At QSI Dimension Lab
Visual Inspection	: OK		

DETAILS OF UNIT UNDER CALIBRATION

Make/ Model	Range	Location	Optical parallel Sr. No.	Series / Model No.	ID No.
MITUTOYO 157-903	Flatness : 0.1 µm Parallelism : 0.2 µm Diameter : 30mm Thickness : (12.0, 12.12, 12.25, 12.37)mm ±0.01mm	-----	021478	157-101	NCQC/M-172
			007579	157.102	
			007531	157-103	
			007740	157-104	

DETAILS OF STANDARD EQUIPMENTS USED FOR CALIBRATION

Name	Make	Sr. No./ID No.	Certificate No.	Validity	Calibrated By
Optical flat	MICRON	100915	MetR1/2021-40	03/12/2023	NABL LAB CC-2006 (CMERI)
Two Probe Comparator (Gauge block calibrator)	KCP-TESA	207 QSI/SGC/01	QSI/2497/21/09	14/09/2023	NABL LAB CC-2717

OBSERVATIONS

DESCRIPTION				Observed Flatness Of face A- Top (µm)	Observed Flatness Of face B- Top (µm)	Observed Parallelism (µm)
Model of set	Series	Sr. No	Thickness in mm			
157-903	157-101	021478	12.00 ± 0.01	0.0492	0.0369	0.02 µm
	157-102	007579	12.12 ± 0.01	0.0369	0.0492	0.04 µm
	157-103	007531	12.25 ± 0.01	0.0614	0.0369	0.05 µm
	157-104	007740	12.37 ± 0.01	0.0492	0.0614	0.03 µm

ULR NO. CC271723000001498F

UUC:- Unit Under Calibration

Format no : F01(QMP-21) -7.8 , Issue no/ Date : 01/01/11/2019 , Revision no/ Date : 00 / 01/11/2019

Traceability of Standard(s) : The Standard Used for calibration is traceable to National Standards

Uncertainty of Measurement (At approximately 95% Confidence Level, k=2) : ± 0.11 µm

Remark : Ref. Standard IS-5440-2020

RESULT: Results of UUC compliance to required Accuracy of Manufacturers (Flatness : 0.1 µm & Parallelism : 0.2 µm) Results OK

Suggested Due Date : 08/09/2025

Calibrated By
Shri Ram Dhiman
(Technical Executive)

Issued By
Ajay Kumar Singh
(Technical Manager)

NCQC

Valid up to 08-09-2025

Reviewed

NCQC System Certificate No. 350

NATIONAL AWARD WINNER