



CC - 2346

TRUE COPY**M.M.D. Kantawala**
CALIBRATION LABORATORY**Calibration Certificate**

Name of Customer → National Centre For Quality Calibration 4, Abhishree Corporate park, Nr.Swagat Bunglows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad – 380058, Gujarat, India.	Certificate No.	MMD/200822/05
	Date of Issue	22-08-2022
	Date of Calibration	20-08-2022
	# Due Date	19-08-2024

Date Of Receipt / Ref. No. → 18-08-2022	F/CAL/02/CR, Issue No.04 Page 1 of 1
--	---

Discipline	Mechanical – Mass	ULR No.	CC2346220000031 4F
Details of Observation of Unit Under Calibration (UUC)	Identification No. : NCQC/M-169	Name of Instrument : Weight	
Weights	50000g/1 Total = 01 pcs		
Type	Cylindrical Handle Type		
Material	Stainless Steel		
Assumed Density	7950 ± 140 kg/m ³		
Make	Weightronics	Visual Inspection	OK
Location	====	Barometric Pressure	1005 ± 15 hPa

Results of Calibration

Id. No.	Denomination (Unit under calibration)	Mass value in g.	Deviation in g.	Uncertainty (±) in mg.	Class
NCQC/M-169	50000 g.	50000.025	0.025	13	E2

- Remarks:**
- Mass values of all the weights are conventional mass values and within the maximum errors permissible in "E2" Accuracy Class of weights as per OIML R 111-1.
 - Thermal stabilization time 48 hours.
 - These results are obtained at the time of calibration.
 - Weights are calibrated for scientific or industrial purpose only.
 - # Due date is given as suggested by customer.
 - Any hand written corrections (except @ marked) or photocopies of the report invalidates this certificate.
 - Environment condition during calibration: 24 ± 0.5°C, 40 to 60% Rh. (Change in temperature and relative humidity during the calibration were less than ± 0.3 °C per hour with a maximum of ± 0.5 °C per 12 hours, and ± 5% Rh per 4 hours respectively)
 - Average temperature → 24.1°C, average pressure 997 hPa and average humidity 47.2% Rh during calibration of instruments.
 - The uncertainties are for a confidence probability of not less than 95.45% with coverage factor k = 2.
 - Condition of instrument found satisfactory during receipt.
 - Calibration is performed on the electronic weighing balance against standards mass by comparison method under controlled conditions (ABBA Method).
 - All calibration performed by MMD Kantawala Calibration Laboratory. None of the results reported in this certificate are from external provider.
 - Calibration results reported in this certificate relates only to the item calibrated.
 - Data provided by customer: Identification No. and accuracy of UUC.
 - Reference standard no.: OIML R 111 – 1 for calibration and classification of weights.
 - Reference Calibration method no.: MMD/CM/02.
 - Master equipment / reference standards are traceable to NABL accredited calibration laboratory.
 - Corrected mass value is calculated based on ABBA method.
 - Calibration was performed using sub-division / sub-multiplication method to calibrate unit under calibration.

Details of Master Instrument Used for Calibration

Nomenclature	Make / Model	Sr. No. / Id. No.	Class	Calibrated by	Calibration certificate no.	Calibration Due Date
Reference Weight Box	MMD / =====	MMD/CAL/02	E1	CC-2854	TC/6944/2020	15-12-2023
Mass Comparator	Mettler Toledo / XPE64003LC	B541536029	II	Not applicable	Not applicable	=====

NCQC
Valid up to 19-08-2024
Reviewed: *[Signature]*

NCQC System Certificate No. 360

Traceable To National / International Standards.

Calibrated By	<i>Vilas</i> Vilas Prajapati Calibration Engineer	Reviewed and Approved By	<i>[Signature]</i> Viral Mistry Technical Manager
----------------------	--	---------------------------------	--