


# Caliber Gauges And Instruments Laboratory LLP



S. No. 14/1, Dattadigamber Colony B, Old Warje Jakat Naka, Pune - 411052.  
+ 91 9422311121 @ caliberlab@yahoo.co.in www.caliberlaboratory.co.in

ULR : CC265625000010937F

PAGE 1 of 1

CERTIFICATE NO : CGIL/2025/0602/H-001

## CALIBRATION CERTIFICATE

TRUE COPY

CALIBRATED FOR

NCQC Laboratory LLP.  
4, Abhishree Corporate Park, Nr. Swagat Bungalow BRTS, Iskon-Ambli  
Road, Ambli, Ahmedabad, Gujarat 380058

TYPE OF INSTRUMENT

Glass Scale

| RANGE      | L.C.   | Identification | Make | Mfg. Sr. No. |
|------------|--------|----------------|------|--------------|
| 0 To 10 mm | 0.1 mm | NCQC/M-115     | -    | 10/100       |

| Challan No. | W. O. No. | Receipt Date | Calibration Date | Due Date (customer decided) | Issue Date |
|-------------|-----------|--------------|------------------|-----------------------------|------------|
| --          | 0602/H    | 02/06/2025   | 03/06/2025       | 02/06/2027                  | 11/06/2025 |

Condition of Item : Ok

Calibration Location : In Lab

Standards / Equipment are traceable to the National Standard

| Description | Identification No. | Certificate No. | Traceability | Valid up to |
|-------------|--------------------|-----------------|--------------|-------------|
| Laser Head  | CGIL/EQP/15        | 111-26124       | NMI          | 06/11/2026  |

CALIBRATION PROCEDURE NO.

: CGIL-CP-35

REFERENCE STANDARD

: JIS B 7541

ENVIRONMENTAL CONDITION

: Temperature : 20.3°C Humidity : 55 %Rh

UNCERTAINTY OF MEASUREMENT

:  $\pm 0.8 \mu\text{m}$  ( At 95.45% Confidence Level i.e.  $k=2$  )

Mechanical Calibration: Dimension (Precision)

## CALIBRATION RESULTS

| Sr. No. | Nominal Value on DUC (in mm) | Observed Value on Master (in mm) | Observed Error (in mm) |
|---------|------------------------------|----------------------------------|------------------------|
| 1       | 0.00                         | 0.0000                           | 0.0000                 |
| 2       | 0.10                         | 0.1003                           | -0.0003                |
| 3       | 0.50                         | 0.5003                           | -0.0003                |
| 4       | 1.00                         | 1.0002                           | -0.0002                |
| 5       | 2.00                         | 1.9998                           | 0.0002                 |
| 6       | 3.00                         | 3.0002                           | -0.0002                |
| 7       | 4.00                         | 3.9998                           | 0.0002                 |
| 8       | 5.00                         | 5.0002                           | -0.0002                |
| 9       | 6.00                         | 6.0004                           | -0.0004                |
| 10      | 7.00                         | 6.9998                           | 0.0002                 |
| 11      | 8.00                         | 8.0002                           | -0.0002                |
| 13      | 9.00                         | 9.0002                           | -0.0002                |
| 14      | 10.00                        | 10.0003                          | -0.0003                |

Calibrated By

NCQC LABORATORY LLP  
Valid up to 02/06/2027  
Reviewed

Technical Assistant  
Atul Sathe



NCQC System Certificate No. 216-1

Approved By

Technical Manager  
Santosh Raskar

- This certificate pertains only to the items calibrated at CGIL Laboratory.
- This report should not be reproduced without a written permission of the laboratory.
- The results reproduced in this certificate are valid under stated conditions at the time of calibration.

x-x-x-x-x- End of certificate -x-x-x-x-x-



CGILF : 7.8:01:00