

Caliber Gauges And Instruments Laboratory LLP

S. No. 14/1, Dattadigamber Colony B, Old Warje Jakat Naka, Pune - 411052.
☎ + 91 9422311121 ✉ caliberlab@yahoo.co.in 🌐 www.caliberlaboratory.co.in



CC-2656

ULR : CC265625000010940F

PAGE 1 of 1

CERTIFICATE NO : CGIL/2025/0602/H-003

CALIBRATION CERTIFICATE

TRUE COPY

CALIBRATED FOR

: NCQC Laboratory LLP.

4, Abhishree Corporate Park, Nr. Swagat Bunglow BRTS, Iskon-Ambli
Road, Ambli, Ahmedabad, Gujarat 380058

TYPE OF INSTRUMENT

: Glass Scale

RANGE	L.C.	Identification	Make	Mfg. Sr. No.
0 To 10 mm	0.1 mm	NCQC/M-17	--	EP-01

Challan No.	W. O. No.	Receipt Date	Calibration Date	Due Date (customer decided)	Issue Date
--	0602/H	02/06/2025	03/06/2025	02/06/2027	11/06/2025

Condition of Item : Ok

Calibration Location : In Lab

Standards / Equipment are traceable to the National Standard

Description	Identification No.	Certificate No.	Traceability	Valid up to
Laser Head	CGIL/EQP/15	111-26124	NMI	06/11/2026

CALIBRATION PROCEDURE NO.

: CGIL-CP-35

REFERENCE STANDARD

: JIS B 7541

ENVIRONMENTAL CONDITION

: Temperature : 20.1°C Humidity : 55 %Rh

UNCERTAINTY OF MEASUREMENT

: $\pm 0.8 \mu\text{m}$ (At 95.45% Confidence Level i.e. $k=2$)

Mechanical Calibration: Dimension (Precision)

CALIBRATION RESULTS

Sr. No.	Nominal Value on DUC (in mm)	Observed Value on Master (in mm)	Observed Error (in mm)
1	0.00	0.0000	0.0000
2	0.10	0.0998	0.0002
3	0.50	0.5002	-0.0002
4	1.00	1.0002	-0.0002
5	2.00	1.9998	0.0002
6	3.00	3.0002	-0.0002
7	4.00	3.9998	0.0002
8	5.00	5.0004	-0.0004
9	6.00	6.0002	-0.0002
10	7.00	7.0005	-0.0005
11	8.00	7.9998	0.0002
13	9.00	8.9996	0.0004
14	10.00	9.9998	0.0002

Calibrated By

ATM

Technical Assistant

Atul Sathe

NCQC LABORATORY LLP

Valid up to 02/06/2027

Reviewed

[Signature]



NCQC System Certificate No. 216

Approved By

[Signature]

Technical Manager

Santosh Raskar

- This certificate pertains only to the items calibrated at CGIL Laboratory.
- This report should not be reproduced without a written permission of the laboratory.
- The results reproduced in this certificate are valid under stated conditions at the time of calibration.

x-x-x-x-x- End of certificate x-x-x-x-x

