



NCQC LABORATORY LLP

4, Abhishree Corporate Park, Nr. Swagat Bunglow BRTS,
 Iskcon-Ambli Road, Ambli, Ahmedabad-380 058 • Ph. +91-79-29795322, 29795323,
 Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
 E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com
 Visit our Web Site : www.calibrationlaboratory.in • LLP IN : ACH-9094 / GSTIN : 24AAWFN8511B1Z9



CC-2128

Precision Calibration Centre with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/M/039/01 Issue No.02

Name of Customer NCQC LABORATORY LLP 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No.	NCQC-M/200924/028
	Date of Issue	20/09/2024
	Date of Calibration	20/09/2024
	Suggested Due Date	19/09/2025
Date of Receipt / Ref.No.: 19/09/2024		

ULR No.: CC21282400006579F	Discipline: Mechanical Calibration Dimension (Basic-Measuring Instrument, Gauges etc.)
-----------------------------------	--

Details of Observation of Unit Under Calibration	Identification No.: NCQC/M-129
	Serial No.: Probe:13944 / DRO: 11273017
	Name of Instrument: Electronic Probe with D.R.O

Results of Calibration


Unit	Reading set on UUC in mm	Reading Observed by Reference Instrument in mm	Error	Expanded Uncertainty in μm ($k=2$) \pm
mm	0.0000	0.0000	0.0000	0.51
mm	0.1000	0.1001	0.0001	0.51
mm	0.2000	0.2002	0.0002	0.51
mm	0.3000	0.3002	0.0002	0.51
mm	0.4000	0.4003	0.0003	0.51
mm	0.5000	0.5004	0.0004	0.51
mm	1.0000	1.0003	0.0003	0.51
mm	1.5000	1.5004	0.0004	0.51
mm	2.0000	2.0005	0.0005	0.51
mm	2.5000	2.5005	0.0005	0.51
mm	3.0000	3.0005	0.0005	0.51
mm	3.5000	3.5006	0.0006	0.51
mm	4.0000	4.0007	0.0007	0.51
mm	4.5000	4.5007	0.0007	0.51
mm	5.0000	5.0008	0.0008	0.51
mm	10.0000	10.0009	0.0009	0.51
mm	15.0000	15.0010	0.0010	0.51
mm	20.0000	20.0011	0.0011	0.51
mm	25.0000	25.0013	0.0013	0.51

Remarks:

- Averages of minimum five readings are reported.
- Suggested due date is given based on customer requirements
- These results are obtained at the time of calibration.
- Condition of instrument found satisfactory during receipt.
- Result relates to the item calibrated only.
- No external provider was used for calibration and hence it is not applicable.
- Calibration certificate shall not be reproduced except in full without written approval of Director, NCQC.
- Any hand written corrections (except @) or photocopies of the report invalidates this certificate.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor $k=2$.
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with National / international standard.

Traceable To National / International Standards.

Calibrated By	Maulik Rathod Calibration Engineer	Reviewed & Approved By	Jigar Pancha Calibration Incharge
----------------------	---------------------------------------	-----------------------------------	--------------------------------------



NCQC System Certificate No. 292

Page 2 of 2

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"



NCQC LABORATORY LLP

4, Abhishree Corporate Park, Nr. Swagat Bunglow BRTS,
 Iskcon-Ambli Road, Ambli, Ahmedabad-380 058 • Ph. +91-79-29795322, 29795323,
 Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
 E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com



CC-2128

Visit our Web Site : www.calibrationlaboratory.in • LLP IN : ACH-9094 / GSTIN : 24AAWFN8511B1Z9

Precision Calibration Centre with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/M/039/01 Issue No.02

Name of Customer NCQC LABORATORY LLP 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No.	NCQC-M/200924/028
	Date of Issue	20/09/2024
	Date of Calibration	20/09/2024
	Suggested Due Date	19/09/2025
	Date of Receipt / Ref.No.: 19/09/2024	

ULR No. : CC212824000006579F	Discipline : Mechanical Calibration Dimension (Basic-Measuring Instrument, Gauges etc.)
------------------------------	--

Details of Observation of Unit Under Calibration	Identification No. : NCQC/M-129
	Serial No. : Probe:13944 / DRO: 11273017
	Name of Instrument : Electronic Probe with D.R.O

Range	0 to 25 mm	Initial Error	Nil
Resolution	0.0001 mm	Visual Inspection	Satisfactory
Make / Model	Sylvac/D50S		
Location	Lab.		

Loc. of Calibration	At Laboratory	Std. Ref. Used	By Comparison Method
Environment Cond.	20 ± 2°C, 40 to 60 %Rh	Cali.Method No.	NCQC/CM/M/039
Actual Env. Cond.	20.8 °C 51.1 %Rh		

Master Instruments Detail

NCQC SYS.NO.	Nomenclature	Make/ Model	Id / Serial No.	Traceable To	Certificate No.	Valid Upto
226	Length Measuring Machine	Octagon / LMM300T	NCQC/M-24 / 050	CC-2751	OM / 46 / 220048	02/03/2025

Traceable To National / International Standards.

Calibrated By	 Maulik Rathod Calibration Engineer	Reviewed & Approved By	 Vignar Panchal Calibration Incharge
---------------	---	------------------------	--



Page 1 of 2