

**Calibration Certificate**

**TRUE COPY**

FF – 24		Page 1	
Certificate No : VIS/24-25/M-0691	ULR No. : CC26952400000691F	Date of Calibration : 22.06.2024	Recommended Due Date : 22.06.2025
Customer Details : National Centre For Quality Calibration 4, Abhishree Corporate Park, Near Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat, India.		Calibrated at : Mechanical Lab	Discipline : Mechanical Calibration
Test Instrument : Digital Pressure Gauge		Serial No. : NVEM1806034	ID No. : NCQC/M-146
Model : MGA N32-2	Range : 0 to 70 bar	Make : Vijay Enterprises	Group : Pressure Indicating Device
Least Count : 0.001 bar	Accuracy : +/- 0.05% FS	Receipt Date : 22.06.2024	Cond. On Receipt : Satisfactory
Details of Standard Used : Hydraulic Dead Weight Tester		Sr. No./Certi.No. : 71760 / GEC/NB/14148-P	Range : 1kg/cm2 to 55kg/cm2
Name/Make : Fluke Calibration Phoenix	Traceability : Gatrada, Ahmedabad.	Valid Up to : 27.11.2025	Standard Used : DKD-R-6-1
Work Instruction No : CP - 01			ISO-GUM-1995
Environment Details		Relative Humidity : (40 - 70)%	
Temperature : (23 +/- 1.5)°C			

**Remarks**

- The reported Expanded Uncertainty is calculated at 95% C.L. with Coverage Factor  $k=2$ .
- Readings are taken in kg/cm2 (PRESSURE) & mmHg (VACUUM)
- CONVERSION :-  
1 kg/cm2 = 0.98067 bar  
1 mmHg = 0.00133 bar
- Test reading is corrected for Local gravity and temperature.
- Local gravity = 9.78763363 m/s<sup>2</sup> Uncertainty = 50 ppm

SIDDHARTH  
VIJAY PATIL

Digitally signed by  
SIDDHARTH VIJAY  
PATIL  
Date: 2024.06.25  
17:30:39 +05'30'

Calibrated By: Harshvardhan Kumar  
Lab Engineer

Approved By:  
Technical Manager / Siddharth Patil (CEO)

NCQC  
Valid up to 22-06-2025  
Reviewed *[Signature]*

NCQC System Certificate No. 299

## Calibration Certificate

Certificate No: VIS/24-25/M-0691

Page 2

ULR No. : CC26952400000691F

### Calibration Results

Sr. No.	Standard DWT Read.	Test Reading						Mean Value	Deviation	Deviation Allowed +/-	Repeatability	Hysteresis	Uncertainty +/-
		Up	Down	Up	Down	Up	Down						
		bar	bar	bar	bar	bar	bar						
1	0.0000	0.000	0.000	0.000	0.000	0.000	0.000	0.0000	0.0350	0.0000	0.0000	0.01291	
2	8.8260	8.828	8.827	8.827	8.826	8.828	8.826	8.8270	0.0010	0.0350	0.0010	0.00231	
3	17.6521	17.655	17.654	17.655	17.654	17.653	17.654	17.6542	0.0021	0.0350	0.0020	0.00253	
4	26.4781	26.482	26.481	26.481	26.483	26.482	26.481	26.4817	0.0036	0.0350	0.0020	0.00344	
5	35.3041	35.308	35.310	35.309	35.309	35.310	35.309	35.3092	0.0051	0.0350	0.0020	0.00459	
6	45.1108	45.118	45.117	45.119	45.118	45.119	45.118	45.1182	0.0074	0.0350	0.0010	0.00586	
7	52.9562	52.965	52.966	52.966	52.965	52.967	52.966	52.9658	0.0096	0.0350	0.0020	0.00688	
8	62.7629	62.774	62.775	62.775	62.774	62.776	62.775	62.7748	0.0119	0.0350	0.0020	0.00816	
9	70.6082	70.621	70.622	70.622	70.621	70.621	70.620	70.6212	0.0130	0.0350	0.0020	0.00918	

All reading are taken in kg/cm2 at the time of calibration and converted to bar.

**SIDDHARTH  
VIJAY PATIL**

Digitally signed by  
SIDDHARTH VIJAY  
PATIL  
Date: 2024.06.25  
17:31:02 +05'30'

Calibrated By: Harshvardhan Kumar  
Lab Engineer

Approved By:  
Technical Manager / Siddharth Patil (CEO)

**Note :**

- The above results are without any adjustment / repair.
- Equipment used for calibration are calibrated & traceable to National & International Standards.
- The calibration results reported are valid at the time of and under stated conditions of the measurements.
- This certificate refers only to particular Items submitted for calibration.
- VIS is not liable for any change in calibration data & performance specification on account of malfunctioning of Standards/Instruments/Equipment covered by this certificate due to damage caused to it after issuance of this certificate.

---X--- END OF REPORT ---X---