



NCQC LABORATORY LLP

4, Abhishree Corporate Park, Nr. Swagat Bunglow BRTS,
Iskcon-Ambli Road, Ambli, Ahmedabad-380 058 • Ph. +91-79-29795322, 29795323,
Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
E-mail : ncqc@calibrationlaboratory.in • calibrationlab.ncqc@gmail.com



Visit our Web Site : www.calibrationlaboratory.in • LLP IN : ACH-9094 / GSTIN : 24AAWFN8511B1Z9

Precision Calibration Centre with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/E/024/02 Issue No.02

Name of Customer

NCQC LABORATORY LLP

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS,
Iskcon - Ambli Road, Ambli, Ahmedabad - 380 058,
Gujarat

Certificate No.

NCQC-E/260525/002

Date of Issue

02/06/2025

Date of Calibration

26/05/2025 to 31/05/2025

Suggested Due Date

30/05/2026

Date of Receipt / Ref.No.: 25/05/2025

ULR No. : CC21282500001882F

Discipline : Electro-Technical Calibration
Time & Frequency

Details of Observation of Unit Under Calibration

Identification No. : NCQC/E-09

Serial No. : ===

Name of Instrument : Digital Stop watch

TRUE COPY

Range 0 to 24 hours

Resolution 0.001 sec

Make / Model Robic/SC-501W

Location Laboratory

Initial Error

Visual Inspection

Nil

Satisfactory

Loc. of Calibration

At Laboratory

Environment Cond.

25 ± 4 °C, 30 to 75 % RH

Actual Env. Cond.

25.6 °C 49.0 %RH

Std. Ref. Used

NIST 960-12

Cali.Method No.

NCQC/CM/E/024

Master Instruments Detail

NCQC SYS.NO.	Nomenclature	Make/ Model	Id / Serial No.	Traceable To	Certificate No.	Valid Upto
374/1	Timer/Stop Watch Calibrator	Digital Instrumentation / =====	NCQC/DTSC-01/ =====	CC-2128	NCQC-E/041024/001	10/10/2025

Calibration Results

Reading observed by UUC in Hr:min:Sec:1/1000s	Corresponding second	Reading observed by Reference in seconds	Absolute Error in second	Expanded Uncertainty in Second (k =2) ±
00:00:01:041	1.041	1.021	0.020	0.04
00:00:20:212	20.212	20.110	0.102	0.12
00:01:00:262	60.262	60.105	0.157	0.06
00:15:00:321	900.321	900.147	0.174	0.15
00:30:00:496	1800.496	1800.288	0.208	0.30
01:00:00:792	3600.792	3600.359	0.433	0.59
05:00:01:212	18001.212	18000.187	1.025	11.69
23:59:51:451	86391.451	86390.320	1.131	25.30

NCQC System Certificate No. 30

Traceable To National / International Standards.

Calibrated By

A. Thombare

Bhagyashri Thombare
Calibration Engineer

Reviewed & Approved By

Vijay Panchal
Calibration Incharge

Page 1 of 2

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"



NCQC LABORATORY LLP

4, Abhishree Corporate Park, Nr. Swagat Bunglow BRTS,
Iskcon-Ambli Road, Ambli, Ahmedabad-380 058 • Ph. +91-79-29795322, 29795323,
Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328816370
E-mail : ncqc@calibrationlaboratory.in • calibrationlab.ncqc@gmail.com



Visit our Web Site : www.calibrationlaboratory.in • LLP IN : ACH-9094 / GSTIN : 24AAWFN8511B1Z9

Precision Calibration Centre with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/E/024/02 Issue No.02

Name of Customer

NCQC LABORATORY LLP

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS,
Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058,
Gujarat

Certificate No.

NCQC-E/260525/002

Date of Issue

02/06/2025

Date of Calibration

26/05/2025 to 31/05/2025

Suggested Due Date

30/05/2026

Date of Receipt / Ref.No.: 25/05/2025

ULR No. : CC212825000001882F

Discipline : Electro-Technical Calibration
Time & Frequency

Details of Observation of Unit Under Calibration

Identification No. : NCQC/E-09

Serial No. : ===

Name of Instrument : Digital Stop watch

Remarks:

- Averages of minimum three readings are reported.
- Suggested due date is given based on customer requirements
- These results are obtained at the time of calibration.
- Condition of instrument found satisfactory during receipt
- Result relates to the item calibrated only.
- No external provider was used for calibration and hence it is not applicable.
- Calibration certificate shall not be reproduced except in full without written approval of Director, NCQC.
- Any hand written corrections (except @) or photocopies of the report invalidates this certificate.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor $k = 2$
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with National / international standard.

Traceable To National / International Standards.

Calibrated By

Bhagyashri Thombare

Bhagyashri Thombare
Calibration Engineer

Reviewed & Approved By

Jibhat Ranchal
Calibration Incharge

*** End of Calibration Certificate ***

Page 2 of 2

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"