



Since 1998

NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058

Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370

E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com

Visit our Web Site : www.calibrationlaboratory.in



CC-2128

TRUE COPY

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/T/010/02 Issue No. 02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No. NCQC-T/290124/006
	Date of Issue 31/01/2024
	Date of Calibration 29/01/2024
	Suggested Due Date 28/01/2025
Date of Receipt / Ref.No.: 26/01/2024	

ULR No. : CC21282400000484F	Discipline : Thermal Calibration Specific Heat and Humidity
------------------------------------	---

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T-158
	Serial No. : 3357
	Name of Instrument : Temperature & Rh Data Logger

Range 0 to 100 % Rh	Visual Inspection Accuracy OK ±3% RH 20 to 80% & Other ±5 % RH
Resolution 0.1 % Rh	
Make / Model Elitech/Rc - 4HC	
Location Laboratory	

Loc. of Calibration At Laboratory	Std. Ref. Used NIST 250-83
Environment Cond. 25 ± 4 °C, 30 to 75 % Rh	
Actual Env. Cond. 25.9 °C 53.0 %Rh	
Cali.Method No. NCQC/CM/T/010	

Master Instruments Detail

NCQC SYS.NO.	Nomenclature	Make/ Model	Id / Serial No.	Traceable To	Certificate No.	Valid Upto
335	Digital thermo Hygro Palm	Rotronic / HP32	NCQC/T-161/ 5181935 / 20281776	CC-2733	30070326	05/07/2024
435	4 wire RTD with Indicator	RTD - Tempens / ====, Indicator - Tempens / TEMPMET-09-02	RTD-NCQC/T-235 / 23000299, Indicator - NCQC/T-236 / T0064	CC-2128	NCQC-T/190923/001	18/09/2024

Calibration Results of Humidity @ 25 °C


Set point on Hybrid Temperature and Humidity Calibrator in % Rh	Reading observed by UUC in % Rh	Reading observed by Master instrument in % Rh	Absolute Error in % Rh	Expanded Uncertainty in % Rh (k =2) ±
20.0	20.7	20.11	0.59	1.85
40.0	41.5	40.17	1.33	1.85
60.0	62.1	60.38	1.72	1.85
80.0	82.7	80.40	2.30	1.85
95.0	98.5	95.48	3.02	1.85

NCQC System Certificate No. 322

Traceable To National / International Standards.

Calibrated By Himanshu Parbadiya
Calibration Engineer

Reviewed & Approved By Vishal Chauhan
Calibration Incharge





Since 1998

NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058

Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370

E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com

Visit our Web Site : www.calibrationlaboratory.in



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/T/010/02 Issue No. 02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No. NCQC-T/290124/006	
	Date of Issue 31/01/2024	
	Date of Calibration 29/01/2024	
	Suggested Due Date 28/01/2025	
Date of Receipt / Ref.No.: 26/01/2024		

ULR No. : CC21282400000484F	Discipline : Thermal Calibration Specific Heat and Humidity
------------------------------------	---

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T-158
	Serial No. : 3357
	Name of Instrument : Temperature & Rh Data Logger

Range -30 to 60 °C	Accuracy ± 0.5 °C
Least Count 0.1 °C	

Calibration Results of Temperature

Set point on Hybrid Temperature and Humidity Calibrator in °C	Reading observed by UUC in °C	Reading observed by Master instrument in °C	Absolute Error in °C	Expanded Uncertainty in °C (k=2) ±
-25.0	-25.5	-25.159	0.341	0.44
-10.0	-10.3	-10.109	0.191	0.44
0.0	-0.1	-0.021	0.079	0.44
10.0	9.9	10.049	0.149	0.44
30.0	29.8	30.073	0.273	0.44
50.0	49.7	50.111	0.411	0.44

Remarks:

- Averages of minimum five readings are reported.
- Suggested due date is given based on customer requirements
- Calibration Points are given based on customer requirements.
- Condition of instrument found satisfactory during receipt.
- These results are obtained at the time of calibration.
- Result relates to the item calibrated only.
- No external provider was used for calibration and hence it is not applicable.
- Calibration certificate shall not be reproduced except in full without written approval of Director, NCQC.
- Any hand written corrections (except @) or photocopies of the report invalidates this certificate.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with National / international standard.
- Portable Hybrid temperature and Humidity calibrator was used as a source of generating temperature and humidity as well as controlling temperature at specified set point.

Traceable To National / International Standards.

Calibrated By Himanshu Parbadiya Calibration Engineer	Reviewed & Approved By Vishal Chauhan Calibration Incharge
--	---

