

NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058 Ph. +91-79-29795322, 29795323 • Fax: +91-79-29795323, Cell No. +91-9327017517, +91-9328616370

*Page 1 of 2

E-mail: ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com

Visit our Web Site: www.calibrationlaboratory.in

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Name of Customer National Centre for		A								
		Calibrat	ion Ce	rtificate					02 Issue No. 02	
National Centre for	Name of Customer					NCQC-T/290124/013				
	National Centre for Quality Calibration					31/01/2024 29/01/2024				
4, Abhishree Corporate Iskcon – Ambli Road, A		Date of Calibration Suggested Due Date 28/01/2025								
Gujarat		Date of Receipt / Ref.No.: 26/01/2024								
ULR No. : CC21282	1000000491	IF Discipline		al Calibration c Heat and Hum	nidity					
Details of Observat	on of Ider	ntification No.	: NCQ	C/T-162					2	
Unit Under Calibrat	on Seri	al No. ne of Instrumen		86H00537 erature & Rh [Data I og	ner				
Range 0 to 100		ic or moduliner	it . remp	l alaic a rai e	Jala Log	901			· ·	
Resolution 0.1 % R				Visual Inspe	ок					
Make / Model Elitech/F				Accuracy		±3% RH 20 to 80% & Other ±5 % RH				
Location Laborat	ory			Me to the						
Loc. of Calibration A	Laboratory	8	•	Std. Ref. Us	sed	ed NIST 250-83				
	± 4 °C, 30 to			Cali.Method	l No	NCQC/CM/T/010				
Actual Env. Cond. 2	5.7 °C	52.7 %Rh		- Call.Wicthoo		11100	0/0141/1/0			
	V	M	aster Ins	truments Det	tail		,			
NCQC Nomer SYS.NO.	clature	Make/ Model	ld	/ Serial No.	Traceal	ble To Certificate No			Valid Upto	
335 Digital therm	Hygro Palm			5181935 / 20281776				70326	05/07/2024	
435 4 wire RTD	4 wire RTD with Indicator RTD - Tempsens ====, Indicator Tempsens / TEMPMET-09-0			23000299, Indicator - NCQC/T-236 / T0064		CC-2128 NCQC-T/		190923/001	18/09/2024	
				ts of Humidit	y @ 25 °	C	•			
				g observed by strument in % Rh		lute Error in % Rh		Expanded Uncertainty in % Rh (k =2) ±		
20			0.6 2				0.49		1.85	
40			40.19		0.81			1.85		
60 61.3 80 82.4			60.		0.94				1.85	
				80.51 95.58		1.89 2.52		1.85		
80		98.1		OF EQ		2.52				

NCQC System Certificate No. 33



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058 Ph. +91-79-29795322, 29795323 • Fax: +91-79-29795323, Cell No. +91-9327017517, +91-9328616370



Calibration Incharge

* Page 2 of 2

E-mail: ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com

Visit our Web Site: www.calibrationlaboratory.in

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

		Calibratio	on Ce	ertificate			F/	/CR/T/010/02	2 Issue No. 02
Name of Customer		Certificate No.	NCQC-T/29	0124/0	013				
National Centre for Quality Calibration				Date of Issue		31/01/2024			
	орте	Date of Calibration 29/01/2024							
4, Abhishree Corporate Parl Iskcon – Ambli Road, Ambli	DK 13,	Suggested Due Date 28/01/2025							
Gujarat	ı	Date of Receipt / Ref.No.: 26/01/2024					1		
ULR No. : CC212824000	000491F	Discipline :		nal Calibration	,				,
	. Identif	ication No.	: NCQ	C/T-162				4	
Details of Observation	of Serial		: EF71	86H00537					
Unit Under Calibration	Name	of Instrument	: Temp	perature & Rh Da	ata Log	ger			
Range	1	-30 to 60 °C		Accurac				± 0.5 °C	
Least Count		0.1 °C							
Loadi Godin		Calibra	tion Re	esults of Tempe	rature				
Set point on Hybrid	Peading obs	erved by UUC	Readi	ng observed by	Abs	olute Error ir	ı °C	Expanded	Uncertainty in
Temperature and Humidity Calibrator in °C		in °C Ma		instrument in °C			°C (k =2) ±		
-25.0	-:	25.5	-25.143		0.357		0.44		
-10.0	-10.4		-10.115		0.285		0.44		
0.0	0.1		-0.024		0.124		0.44		
10.0	10.2		10.035		0.165		0.44		
30.0	30.3		30.050		0.250		0.44		
50.0	50.4			50.066		0.334		0.44	
 Suggested due date is give Calibration Points are give Condition of instrument for These results are obtained Result relates to the item of No external provider was to Calibration certificate shall Any hand written correctio The uncertainties are for a Our masters are directly international standard. Portable Hybrid temperature at s 	en based or und satisfa d at the tim calibrated or used for ca I not be rep in confidence calibrated ure and Hi	n customer requictory during rece of calibration. only. libration and he orroduced excep @) or photocope probability of through NAB	ence it is t in full v pies of t not less L accre	s not applicable. without written app he report invalidat than 95% with co dited calibration l	es triis verage aborato	factor k = 2 ory having d	irect t		
		• •							
		3							
	٧,								
								THE FOR Q	UALITA
		Traccable T	o Natio	nal / Internationa	Stand	lards.	1	CALIBRAT	ION
	-6	Himanshu					1	3V 1919	shal Chauha
Calibrated By	74	/ / miniarishu	aibaui	ya Kevieweu	~ whi	u,	/\	\\dibo	ation Inchar

Calibration Engineer