



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
 Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
 E-mail : ncqc@calibrationlaboratory.in • calibrationlab.ncqc@gmail.com



CC-2128

Since 1996

Visit our Web Site : www.calibrationlaboratory.in

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/11/010/02 Issue No. 02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No.	NCQC-T/290124/010
	Date of Issue	31/01/2024
	Date of Calibration	29/01/2024
	Suggested Due Date	28/01/2025
Date of Receipt / Ref.No.:		26/01/2024

ULR No. : CC212824000000488F	Discipline : Thermal Calibration Specific Heat and Humidity
------------------------------	--

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T-164
	Serial No. : EF7186H00539
	Name of Instrument : Temperature & Rh Data Logger

Range	0 to 100 % Rh	Visual Inspection Accuracy	OK ±3% RH 20 to 80% & Other ±5 % RH
Resolution	0.1 % Rh		
Make / Model	Elitech/Rc - 4HC		
Location	Laboratory		

Loc. of Calibration	At Laboratory	Std. Ref. Used	NIST 250-83
Environment Cond.	25 ± 4 °C, 30 to 75 % Rh	Cali.Method No.	NCQC/CM/T/010
Actual Env. Cond.	25.6 °C 51.0 %Rh		

Master Instruments Detail

NCQC SYS.NO.	Nomenclature	Make/ Model	Id / Serial No.	Traceable To	Certificate No.	Valid Upto
335	Digital thermo Hygro Palm	Rotronic / HP32	NCQC/T-161/ 5181935 / 20281776	CC-2733	30070326	05/07/2024
435	4 wire RTD with Indicator	RTD - Tempens / ====, Indicator - Tempens / TEMPMET-09-02	RTD-NCQC/T-235/ 23000299, Indicator - NCQC/T-236 / T0064	CC-2128	NCQC-T/190923/001	18/09/2024

Calibration Results of Humidity @ 25 °C

Set point on Hybrid Temperature and Humidity Calibrator in % Rh	Reading observed by UUC in % Rh	Reading observed by Master instrument in % Rh	Absolute Error in % Rh	Expanded Uncertainty in % Rh (k =2) ±
20	20.6	20.11	0.49	1.85
40	41.0	40.15	0.85	1.85
60	61.2	60.23	0.97	1.85
80	81.9	80.34	1.56	1.85
95	97.6	95.59	2.01	1.85

Traceable To National / International Standards.

Calibrated By

Himanshu Parbadiya
Calibration Engineer

Reviewed & Approved By

Vishal Chauhan
Calibration Incharge

Page 1 of 2

NCQC System Certificate No. 339

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058

Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370

E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com

Visit our Web Site : www.calibrationlaboratory.in



CC-2128

Since 1998

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/T/010/02 Issue No. 02

Name of Customer

National Centre for Quality Calibration

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat

Certificate No.	NCQC-T/290124/010
Date of Issue	31/01/2024
Date of Calibration	29/01/2024
Suggested Due Date	28/01/2025
Date of Receipt / Ref.No.:	26/01/2024

ULR No. : CC21282400000488F

Discipline : Thermal Calibration
Specific Heat and Humidity

Details of Observation of Unit Under Calibration

Identification No. : NCQC/T-164
 Serial No. : EF7186H00539
 Name of Instrument : Temperature & Rh Data Logger

Range	-30 to 60 °C	Accuracy	± 0.5 °C
Least Count	0.1 °C		

Calibration Results of Temperature

Set point on Hybrid Temperature and Humidity Calibrator in °C	Reading observed by UUC in °C	Reading observed by Master instrument in °C	Absolute Error in °C	Expanded Uncertainty in °C (k = 2) ±
-25.0	-25.6	-25.185	0.415	0.44
-10.0	-10.3	-10.123	0.177	0.44
0.0	0.2	-0.029	0.229	0.44
10.0	10.2	10.055	0.145	0.44
30.0	30.4	30.068	0.332	0.44
50.0	50.5	50.087	0.413	0.44

Remarks:

- Averages of minimum five readings are reported.
- Suggested due date is given based on customer requirements
- Calibration Points are given based on customer requirements.
- Condition of instrument found satisfactory during receipt.
- These results are obtained at the time of calibration.
- Result relates to the item calibrated only.
- No external provider was used for calibration and hence it is not applicable.
- Calibration certificate shall not be reproduced except in full without written approval of Director, NCQC.
- Any hand written corrections (except @) or photocopies of the report invalidates this certificate.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with National / international standard.
- Portable Hybrid temperature and Humidity calibrator was used as a source of generating temperature and humidity as well as controlling temperature at specified set point.

Traceable To National / International Standards.

Calibrated By

Himanshu Parbadiya
Calibration Engineer

Reviewed & Approved By



Page 2 of 2