





NCQC LABORATORY LLP

4, Abhishree Corporate Park, Nr. Swagat Bungalow BRTS,
 Iskcon-Ambli Road, Ambli, Ahmedabad-380 058 • Ph. +91-79-29795322, 29795323,
 Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
 E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com



Visit our Web Site : www.calibrationlaboratory.in • LLP IN : ACH-9094 / GSTIN : 24AAWFN8511B1Z9

Precision Calibration Centre with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate				F/CR/E/024/03 Issue No.02		
Name of Customer NCQC LABORATORY LLP 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat		Certificate No. Date of Issue Date of Calibration Suggested Due Date	NCQC-E/041024/001 11/10/2024 04/10/2024 to 11/10/2024 10/10/2025	Date of Receipt / Ref.No.: 04/10/2024		
ULR No. : CC212824000007118F		Discipline : Electro-Technical Calibration Time & Frequency				
Details of Observation of Unit Under Calibration		Identification No. : NCQC/DTSC-01 Serial No. : ===== Name of Instrument : Digital Timer/Stop watch Calibrator				
Range	0 to 99999.999 Sec	Initial Error	Nil			
Resolution	0.001 Sec	Visual Inspection	OK			
Make / Model	Digital Instrumentation/=====	Accuracy	As per Manual			
Location	Laboratory	Type / Size	Digital / ===			
Loc. of Calibration	At Laboratory	Std. Ref. Used	NIST 960–12			
Environment Cond.	25 ± 5°C, 30 to 75% Rh.	Cali.Method No.	NCQC/CM/E/024			
Actual Env. Cond.	25.9 °C 47.7 %Rh					
Master Instruments Detail						
NCQC SYS.NO.	Nomenclature	Make/ Model	Id / Serial No.	Traceable To	Certificate No.	Valid Upto
374	Timer / Stop watch Calibrator	Digital Instrumentation / =====	NCQC/E-29 / DITC080519	CC-2287	CC/ECL/1586/23-24	11/03/2025
Calibration Results						
Set Points on UUC	Unit	Reading observed by Reference in seconds	Absolute Error in second	Expanded Uncertainty in Second (k =2) ±		
0.010	Sec	0.010	0.000	0.002		
60.000	Sec	60.000	0.000	0.035		
900.000	Sec	899.995	0.005	0.104		
1800.000	Sec	1799.988	0.012	0.210		
3600.000	Sec	3599.843	0.157	0.419		
43200.000	Sec	43199.592	0.408	11.507		
86400.000	Sec	86398.518	1.482	23.250		
Traceable To National / International Standards.						
Calibrated By  Krupesh Gajera Calibration Engineer		Reviewed & Approved By  Hitesh Limbachiya Calibration Incharge				







NCQC LABORATORY LLP

4, Abhishree Corporate Park, Nr. Swagat Bunglow BRTS,
 Iskcon-Ambli Road, Ambli, Ahmedabad-380 058 • Ph. +91-79-29795322, 29795323,
 Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
 E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com



Visit our Web Site : www.calibrationlaboratory.in • LLP IN : ACH-9094 / GSTIN : 24AAWFN8511B1Z9

Precision Calibration Centre with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate		F/CR/E/024/03 Issue No.02	
Name of Customer NCQC LABORATORY LLP 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat		Certificate No. Date of Issue Date of Calibration Suggested Due Date	NCQC-E/041024/001 11/10/2024 04/10/2024 to 11/10/2024 10/10/2025
ULR No. : CC212824000007118F		Discipline : Electro-Technical Calibration Time & Frequency	
Details of Observation of Unit Under Calibration	Identification No. : NCQC/DTSC-01 Serial No. : ===== Name of Instrument : Digital Timer/Stop watch Calibrator		
Remarks: <ul style="list-style-type: none"> Averages of minimum three readings are reported. Suggested due date is given based on customer requirements These results are obtained at the time of calibration. Condition of instrument found satisfactory during receipt Result relates to the item calibrated only. No external provider was used for calibration and hence it is not applicable. Calibration certificate shall not be reproduced except in full without written approval of Director, NCQC. Any hand written corrections (except @) or photocopies of the report invalidates this certificate. The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2 Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with National / international standard. 			
Traceable To National / International Standards.			
Calibrated By  Krupesh Gajera Calibration Engineer	Reviewed & Approved By  Hitesh Limbachiya Calibration Incharge		