

TRUE COPY



FINIKS ENGINEERING COMPANY

Plot No. 21, Anand Nagar, Near Wasan Automobiles, Sahyadri Chowk, MIDC, Anagar - 414111, [M.S.] Tel [0241] 2778918, E-mail finiks97@rediffmail.com

Calibration Certificate

ULR : CC2267 23000007262F

CALIBRATE D ON	ISSUED ON	CERTIFICATE NO.	NEXT CALIBRATION DUE ON (suggested by customer)	PAGE	NO. OF PAGES
13/03/2023	13/03/2023	23-MCC-13-12-001	12/03/2024	1	2

CALIBRATED FOR : M/s National Centre For Quality Calibration, 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad. 380058

CUSTOMER'S CHALLAN NO. : 060323/009 Date of Receipt : 13/03/2023 Dated : 06/03/2023 Work Order No. : 23-MC-13-12

TYPE OF GAUGE : THREAD MEASURING WIRE

IDENTIFICATION : MAKE : Aditya Sr. No. : 00-003 SIZE : 0.17 to 3.20 mm ID. No. : NCQC/M-10 Accuracy : ±0.5 µm

LAB ID NO. : 23-MC-13-12-001 ENVIRONMENTAL CONDITION : TEMPERATURE : 20°C (± 1°C) HUMIDITY : 50 ± 10%RH REFERENCE DOCUMENT USED : IS 6311-1978 CONDITION AT RECEIPT : Ok

CALIBRATION PROCEDURE : CP/34/00 CALIBRATION LOCATION : In Lab

Equipment / Reference Standards Used For Calibration : Traceable to National / International standards.

Equipment / Reference Standard	Identification No.	Certificate No.	Calib. Date	Valid up to
Length Measuring Machine	Finiks/LMM-01	CGIL/2022/1121/K-001	21/11/2022	20/11/2024
A SET OF SLIP GAUGE (CARBIDE)	TC-M122/1	KCP/01/21-22/3536	04/08/2021	04/08/2023

UNCERTAINTY OF MEASUREMENT : ± 0.67 µm

The reported uncertainty is at coverage factor K=2 which corresponds to a coverage probability of approximately 95% for normal distribution.


a) CALIBRATION RESULTS ARE ENCLOSED ON PAGE NO.2 ONWARDS.

NCQC Valid up to 12-03-2024 Reviewed [Signature]

NCQC System Certificate No. 38

Calibrated by [Signature] Authorized Signatory [Signature]

This Certificate refers only to the item (s) submitted for calibration. This calibration results in this certificate are valid at the stated conditions of measurements. This certificate shall not be reproduced, except in full, without the permission in writing.



FN/25/00

TRUE COPY



FINIKS ENGINEERING COMPANY

Plot No. 21, Anand Nagar, Near Wasan Automobiles, Sahyadri Chowk, MIDC, Anand Nagar - 414111, [M.S.] Tel: [0241] 2778918, E-mail: finiks97@rediffmail.com

Calibration Certificate

ULR : CC226723000007262F

CALIBRATE D.ON	ISSUED ON	CERTIFICATE NO.	NEXT CALIBRATION DUE ON (suggested by customer)	PAGE	NO. OF PAGES
13/03/2023	13/03/2023	23-MCC-13-12-001	12/03/2024	2	2

Dimension (Basic-Measuring Instrument)

CALIBRATION RESULTS


All reading in mm

Sr. No.	Nominal Size in mm	Observed size in mm (Pin No.A)	Observed size in mm (Pin No.B)	Observed size in mm (Pin No.C)
1	0.170	0.1705	0.1702	0.1703
2	0.195	0.1954	0.1951	0.1953
3	0.220	0.2202	0.2203	0.2204
4	0.250	0.2504	0.2502	0.2502
5	0.290	0.2901	0.2901	0.2899
6	0.335	0.3351	0.3353	0.3349
7	0.390	0.3902	0.3900	0.3903
8	0.455	0.4552	0.4551	0.4550
9	0.530	0.5301	0.5304	0.5302
10	0.620	0.6202	0.6203	0.6201
11	0.725	0.7254	0.7251	0.7252
12	0.895	0.8946	0.8948	0.8947
13	1.100	1.0996	1.1001	1.0999
14	1.350	1.3496	1.3499	1.3498
15	1.650	1.6495	1.6496	1.6495
16	2.050	2.0503	2.0500	2.0502
17	2.550	2.5502	2.5501	2.5503
18	3.200	3.1995	3.2002	3.1998

x-x-x-x-x- End of certificate -x-x-x-x-x-x

Calibrated by Smokal Authorized Signatory Sharma

This Certificate refers only to the item (s) submitted for calibration. This calibration results in this certificate are valid at the stated conditions at measurements. This certificate shall not be reproduced, except in full, without the permission in writing.



FLV25/00