



CALIBRATION CERTIFICATE

Godrej & Boyce Mfg. Co. Ltd.,
Calibration Services, Plant 18 B, 1st Floor,
Pirojshanagar, Vikhroli,
Mumbai - 400079.
Tel.: +91-22-67965914-48

FR-QT-21		ULR No.: CC26342400000611F	
Certificate No. :	M-240109-1-1	Date of Issue :	10-Jan-24
Date of Calibration :	09-Jan-24	Recom. Due Date :	09-Jan-25
Customer Details :			
NATIONAL CENTRE FOR QUALITY CALIBRATION 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, AMBLI ROAD AMBLI AHMEDABAD Ahmedabad Gujarat India-380001.		Calibrated at :	Lab
		Date of Receipt :	09-Jan-24
		Cond. On Receipt :	Satisfactory
		Receipt No. :	M-240109-1
Details of Under Calibration instrument			
Name :	6.5 Digit Multimeter	Make :	Fluke
Model :	8846A	Sr. No. :	1086005
I.D. No. :	NCQC/E-04	Range :	As Per Report
L.C. :	As Per Report		
Details of Master Equipment			
Name	ID No./Sr. No.	Valid Upto	Traceability
Reference Multimeter(8508A)	LQE090/221965838	05-Dec-24	ETDC, Bangalore
Multifunction Calibrator(5720A)	LQE091/2216202	05-Dec-24	ETDC, Bangalore
Multi-Product Calibrator(5522A)	LQE311/5949905	19-Oct-24	Godrej & Boyce, Mumbai
Standard Resistor (1GΩ)	LQE129/221966171	21-Nov-24	ETDC, Bangalore
Work Instruction No. :	WI-ET-04,15,18,20		
Environmental Details :			
Temperature : 24.1°C		Relative Humidity : 57%RH	
REMARK :			
1. Calibration Points are taken as per Customer's Request. 2. Discipline : Electro-Technical, Group : Direct Current, Alternating Current < 1 GHz, Time & Frequency, Temperature Simulation.			
AUTHORIZED SIGNATORY : Asmita Parab DESIGNATION : Sr. Calibration Officer CALIBRATED BY : Lab Engineer : Ravi R Poojary			

NCQC
Valid up to 09/01/2025
Reviewed *[Signature]*




Validity unknown
Digitally signed by DS/M/S GODREJ AND BOYCE MFG CO LTD
Date: 2024.01.10 16:36:17 +05:30
Reason: Calibration Certificate
Location: Mumbai

Godrej & Boyce Mfg. Co. Ltd.

Regd. Office: Pirojshanagar, L.B.S. Marg, Vikhroli, Mumbai - 400 079.T.:+91-22-67961700/1800 W:www.godrej.com
(This document is digitally signed on first page and does not require signature)

TRUE COPY

NCQC System Certificate No. **43**



CALIBRATION CERTIFICATE

Godrej & Boyce Mfg. Co. Ltd.
 Calibration Services, Plant 18 B, 1st Floor,
 Pirojshahagar, Vikhroli,
 Mumbai - 400079.
 Dir.: +91-22-67965914-88

FR-QT-21
 Certificate No.:M-240109-1-1


ULR No.:CC2634-4000006111

Range		UUC Reading	Standard Reading	Error in Unit	Measurement Uncertainty (±)	Coverage Factor @ C.I. 95.45 (k)
DC Voltage (Measure)	100mV	-99.9989 mV	-100.00000 mV	0.00110	0.0019 %	2
		0.9999 mV	1.00000 mV	-0.00010	0.0475 %	2
		10.0000 mV	10.00000 mV	0.00000	0.0056 %	2
		50.0012 mV	50.00000 mV	0.00120	0.0019 %	2
		100.0024 mV	100.00000 mV	0.00240	0.0019 %	2
DC Voltage (Measure)	1V	-0.999999 V	-1.0000000 V	0.0000010	0.0008 %	2
		0.100001 V	0.10000000 V	0.00000100	0.0019 %	2
		0.500002 V	0.5000000 V	0.0000020	0.0009 %	2
		1.000002 V	1.0000000 V	0.0000020	0.0008 %	2
DC Voltage (Measure)	10V	-10.00002 V	-10.000000 V	-0.000020	0.0007 %	2
		1.00000 V	1.0000000 V	0.0000000	0.0010 %	2
		5.00001 V	5.000000 V	0.000010	0.0007 %	2
		10.00003 V	10.000000 V	-0.000030	0.0007 %	2
DC Voltage (Measure)	100V	-100.0002 V	-100.00000 V	-0.00020	0.0008 %	2
		9.9999 V	10.000000 V	-0.000100	0.0008 %	2
		49.9999 V	50.00000 V	-0.00010	0.0008 %	2
		100.0000 V	100.00000 V	0.00000	0.0008 %	2
DC Voltage (Measure)	1000V	-1,000.002 V	-1,000.0000 V	-0.0020	0.0009 %	2
		100.000 V	100.00000 V	0.00000	0.0010 %	2
		500.000 V	500.0000 V	0.0000	0.0010 %	2
		1,000.002 V	1,000.0000 V	0.0020	0.0009 %	2
AC Voltage (Measure)	100mV @ 50Hz	1.0045 mV	1.000000 mV	0.004500	0.4760 %	2
		9.9988 mV	10.00000 mV	-0.00120	0.0579 %	2
		49.9937 mV	50.0000 mV	-0.0063	0.0272 %	2
	100mV @ 1kHz	99.9905 mV	100.0000 mV	-0.0095	0.0200 %	2
		99.9916 mV	100.0000 mV	-0.0084	0.0200 %	2
AC Voltage (Measure)	1V @ 50Hz	0.999962 V	0.1000000 V	-0.0000380	0.0200 %	2
		0.499944 V	0.500000 V	-0.000056	0.0086 %	2
	1V @ 1kHz	0.999905 V	1.000000 V	-0.000095	0.0079 %	2
AC Voltage (Measure)	10V @ 50Hz	0.999911 V	1.000000 V	-0.000089	0.0079 %	2
		0.99966 V	1.000000 V	-0.000340	0.0079 %	2
	4.99924 V	5.00000 V	-0.00076	0.0090 %	2	
	10V @ 1kHz	9.99845 V	10.00000 V	-0.00155	0.0086 %	2
		9.99851 V	10.00000 V	-0.00149	0.0086 %	2

Godrej & Boyce Mfg. Co. Ltd.

Regd. Office: Pirojshahagar, L.B.S. Marg, Vikhroli, Mumbai - 400 079.T.:+91-22-67961700/1800 W:www.godrej.com
 (This document is digitally signed on first page and does not require signature)

TRUE COPY



CALIBRATION CERTIFICATE

Godrej & Boyce Mfg. Co. Ltd.
 Calibration Services, Plant 13 B, 1st Floor,
 Pirojshanagar, Vikhroli,
 Mumbai - 400079,
 Dir: +91-22-67965914-48

FR-QT-21
 Certificate No.:M-240109-1-1

ULR No.:CC2634-400000611F


Range		UUC Reading	Standard Reading	Error in Unit	Measurement Uncertainty (±)	Coverage Factor @ C.I. 95.45 (k)
AC Voltage (Measure)	100V @ 50Hz	9.9966 V	10.00000 V	-0.00340	0.0086 %	2
		49.9895 V	50.00000 V	-0.0105	0.0098 %	2
	100V @ 1kHz	99.9805 V	100.00000 V	-0.0195	0.0092 %	2
		99.9819 V	100.00000 V	-0.0181	0.0092 %	2
AC Voltage (Measure)	1000V @ 50Hz	99.968 V	100.00000 V	-0.0320	0.0093 %	2
		499.919 V	500.00000 V	-0.081	0.0166 %	2
	1000V @ 1kHz	999.772 V	1,000.00000 V	-0.228	0.0164 %	2
		999.807 V	1,000.00000 V	-0.193	0.0164 %	2
DC Current (Measure)	100µA	-100.0059 µA	-100.00000 µA	-0.0059	0.0117 %	2
		10.0068 µA	10.00000 µA	0.0068	0.0900 %	2
		50.0054 µA	50.00000 µA	0.0054	0.0186 %	2
		100.0005 µA	100.00000 µA	0.0005	0.0117 %	2
DC Current (Measure)	1mA	-1.000016 mA	-1.0000000 mA	-0.000016	0.0050 %	2
		0.100009 mA	0.1000000 mA	0.0000090	0.0117 %	2
		0.500023 mA	0.5000000 mA	0.000023	0.0058 %	2
		1.000027 mA	1.0000000 mA	0.000027	0.0050 %	2
DC Current (Measure)	10mA	-9.99752 mA	-10.000000 mA	0.00248	0.0047 %	2
		1.00051 mA	1.0000000 mA	0.000510	0.0050 %	2
		4.99972 mA	5.0000000 mA	-0.00028	0.0051 %	2
		9.99848 mA	10.0000000 mA	-0.00152	0.0047 %	2
DC Current (Measure)	100mA	-99.9762 mA	-100.000000 mA	0.0238	0.0062 %	2
		9.9982 mA	10.0000000 mA	-0.00180	0.0047 %	2
		49.9888 mA	50.0000000 mA	-0.0112	0.0070 %	2
		99.9763 mA	100.0000000 mA	-0.0237	0.0062 %	2
DC Current (Measure)	400mA	-399.894 mA	-400.000000 mA	0.106	0.0129 %	2
		39.989 mA	40.0000000 mA	-0.0110	0.0075 %	2
		199.954 mA	200.0000000 mA	-0.0460	0.0060 %	2
		399.893 mA	400.0000000 mA	-0.107	0.0129 %	2
DC Current (Measure)	1A	-1.000023 A	-1.0000000 A	-0.000023	0.0109 %	2
		0.100048 A	0.10000000 A	0.0000480	0.0062 %	2
		0.500052 A	0.50000000 A	0.000052	0.0122 %	2
		1.000028 A	1.00000000 A	0.000028	0.0109 %	2
DC Current (Measure)	3A	-2.99991 A	-3.0000000 A	0.00009	0.0101 %	2
		0.30007 A	0.30000000 A	0.000070	0.0607 %	2
		1.50006 A	1.50000000 A	0.00006	0.0760 %	2
		2.99967 A	3.00000000 A	-0.00033	0.0101 %	2
DC Current (Measure)	10A	-9.99906 A	-10.0000000 A	0.00094	0.0895 %	2
		1.00004 A	1.00000000 A	0.00004	0.0770 %	2
		4.99943 A	5.00000000 A	-0.00057	0.0944 %	2
		9.99932 A	10.00000000 A	-0.00068	0.0895 %	2

Godrej

Godrej & Boyce Mfg. Co. Ltd.

Regd. Office: Pirojshanagar, L.B.S. Marg, Vikhroli, Mumbai - 400 079.T.:+91-22-67961700/1800 W:www.godrej.com
 (This document is digitally signed on first page and does not require signature)

TRUE COPY



CALIBRATION CERTIFICATE

Godrej CALIBRATION SERVICES
Godrej & Boyce Mfg. Co. Ltd.
 Calibration Services, Plant 18 B, 1st Floor,
 Pirojshanagar, Vikhroli,
 Mumbai - 400079.
 Dir: +91-22-67965914-48

FR-QT-21

Certificate No.:M-240109-1-1


ULR No.:CC26342400000611F

Range		UUC Reading	Standard Reading	Error in Unit	Measurement Uncertainty (±)	Coverage Factor @ C.L. 95.45 (k)
AC Current (Measure)	100µA @ 50Hz	10.0082 µA	10.000 µA	0.0082	0.1100 %	2
		50.0201 µA	50.000 µA	0.0201	0.5910 %	2
	100µA @ 1kHz	100.0436 µA	100.000 µA	0.0436	0.5910 %	2
AC Current (Measure)	1mA @ 50Hz	100.0465 µA	100.000 µA	0.0465	0.5910 %	2
		0.099981 mA	0.100000 mA	-0.000019	0.5910 %	2
	1mA @ 1kHz	0.499994 mA	0.500000 mA	-0.000006	0.0600 %	2
AC Current (Measure)	10mA @ 50Hz	0.999994 mA	1.000000 mA	-0.000006	0.0600 %	2
		0.999982 mA	1.000000 mA	-0.000018	0.0600 %	2
	10mA @ 1kHz	9.99998 mA	10.00000 mA	-0.00002	0.0600 %	2
AC Current (Measure)	100mA @ 50Hz	4.99996 mA	5.00000 mA	-0.00004	0.0600 %	2
		9.99767 mA	10.00000 mA	-0.00233	0.0600 %	2
	100mA @ 1kHz	9.999714 mA	10.00000 mA	-0.000286	0.0600 %	2
AC Current (Measure)	400mA @ 50Hz	9.9955 mA	10.00000 mA	-0.0045	0.0600 %	2
		49.9856 mA	50.00000 mA	-0.0144	0.0600 %	2
	400mA @ 1kHz	99.9716 mA	100.00000 mA	-0.0284	0.0600 %	2
AC Current (Measure)	1A @ 50Hz	99.9695 mA	100.00000 mA	-0.0305	0.0600 %	2
		39.973 mA	40.00000 mA	-0.027	0.0600 %	2
	400mA @ 1kHz	199.888 mA	200.00000 mA	-0.112	0.0600 %	2
AC Current (Measure)	1A @ 1kHz	399.803 mA	400.00000 mA	-0.197	0.0453 %	2
		399.813 mA	400.00000 mA	-0.187	0.0453 %	2
	1A @ 50Hz	0.099991 A	0.100000 A	-0.000009	0.0600 %	2
AC Current (Measure)	3A @ 50Hz	0.499942 A	0.500000 A	-0.000058	0.0436 %	2
		1.000123 A	1.000000 A	0.000123	0.0401 %	2
	3A @ 1kHz	1.000012 A	1.000000 A	0.000012	0.0401 %	2
AC Current (Measure)	10A @ 50Hz	0.29986 A	0.300000 A	-0.000140	0.0595 %	2
		1.49975 A	1.500000 A	-0.00025	0.0810 %	2
	10A @ 1kHz	3.00017 A	3.000000 A	0.00017	0.1510 %	2
AC Current (Measure)	10A @ 50Hz	3.00592 A	3.000000 A	0.00592	0.1960 %	2
		0.99991 A	1.000000 A	-0.00009	0.0737 %	2
	10A @ 1kHz	5.00014 A	5.000000 A	0.00014	0.1210 %	2
RTD Pt 385, 100 Ω Measure	-200°C to 800°C	10.00002 A	10.00000 A	0.00002	0.1100 %	2
		10.02027 A	10.00000 A	0.02027	0.1440 %	2
		-199.995 °C	-200.000 °C	0.005	0.2500 °C	2
		-0.002 °C	0.000 °C	-0.002	0.2500 °C	2
		24.996 °C	25.000 °C	-0.004	0.2500 °C	2
99.998 °C	100.000 °C	-0.002	0.2500 °C	2		
499.994 °C	500.000 °C	-0.006	0.2500 °C	2		
798.996 °C	799.000 °C	-0.004	0.2710 °C	2		

Godrej
Godrej & Boyce Mfg. Co. Ltd.

Regd. Office: Pirojshanagar, L.B.S. Marg, Vikhroli, Mumbai - 400 079.T.:+91-22-67961700/1800 W:www.godrej.com
 (This document is digitally signed on first page and does not require signature)

TRUE COPY



CALIBRATION CERTIFICATE

Godrej & Boyce Mfg. Co. Ltd.,
 Calibration Services, Plant 18 B, 1st Floor,
 Pirojshanagar, Vikhroli,
 Mumbai - 400079.
 Dir.: +91-22-67965914-48

FR-QT-21
 Certificate No.:M-240109-1-1


ULR No.:CC26342400000611F

Range	UUC Reading	Standard Reading	Error in Unit	Measurement Uncertainty (±)	Coverage Factor @ C.I. 95.45 (k)
Frequency (Measure) 10Hz to 1000kHz	10.00001 Hz	10.00 Hz	0.00001	0.1200 %	2
	40.0000 Hz	40.00 Hz	0.0000	0.1200 %	2
	49.9999 Hz	50.00 Hz	-0.0001	0.1200 %	2
	99.9999 Hz	100.00 Hz	-0.0001	0.1200 %	2
	499.999 Hz	500.0 Hz	-0.001	0.1200 %	2
	1.000001 kHz	1.0000 kHz	0.000001	0.1200 %	2
	2.99998 kHz	3.0000 kHz	-0.00002	0.1200 %	2
	4.99997 kHz	5.0000 kHz	-0.00003	0.1200 %	2
	9.99995 kHz	10.0000 kHz	-0.00005	0.1200 %	2
	99.9995 kHz	100.00 kHz	-0.0005	0.1200 %	2
	499.997 kHz	500.0 kHz	-0.003	0.1200 %	2
	999.995 kHz	1,000.0 kHz	-0.005	0.1200 %	2
Resistance (Measure) 10Ω - 4 wire	0.99994 Ω	0.9998171 Ω	0.0001229	0.0120 %	2
	1.89982 Ω	1.8999518 Ω	-0.0001318	0.0111 %	2
	9.99929 Ω	9.997855 Ω	0.001435	0.0030 %	2
Resistance (Measure) 100Ω - 4 wire	9.9950 Ω	9.997855 Ω	-0.002855	0.0030 %	2
	18.9954 Ω	18.996737 Ω	-0.001337	0.0030 %	2
	100.0035 Ω	99.99618 Ω	0.00732	0.0013 %	2
Resistance (Measure) 1kΩ - 4 wire	0.100004 kΩ	0.09999618 kΩ	0.00000782	0.0014 %	2
	0.189999 kΩ	0.18999344 kΩ	0.00000556	0.0013 %	2
	0.999999 kΩ	0.9999774 kΩ	0.0000216	0.0011 %	2
Resistance (Measure) 10kΩ - 4 wire	0.99999 kΩ	0.9999774 kΩ	0.0000126	0.0013 %	2
	1.89997 kΩ	1.8999574 kΩ	0.0000126	0.0012 %	2
	9.99986 kΩ	9.999642 kΩ	0.000218	0.0011 %	2
Resistance (Measure) 100kΩ - 4 wire	9.9998 kΩ	9.999642 kΩ	0.000158	0.0013 %	2
	18.9988 kΩ	18.998503 kΩ	0.000297	0.0012 %	2
	100.0018 kΩ	100.00065 kΩ	0.00115	0.0014 %	2
Resistance (Measure) 1MΩ - 4 wire	0.100002 MΩ	0.10000065 MΩ	0.00000135	0.0015 %	2
	0.189987 MΩ	0.18998361 MΩ	0.00000339	0.0014 %	2
	0.999962 MΩ	0.9999582 MΩ	0.0000038	0.0030 %	2
Resistance (Measure) 10MΩ - 4 wire	1.00000 MΩ	0.9999582 MΩ	0.0000418	0.0030 %	2
	1.89996 MΩ	1.8998802 MΩ	0.0000798	0.0030 %	2
	9.99903 MΩ	9.999065 MΩ	-0.000035	0.0050 %	2
Resistance (Measure) 100MΩ - 4 wire	10.0003 MΩ	9.999065 MΩ	0.001235	0.0050 %	2
	19.0024 MΩ	19.000109 MΩ	0.002291	0.0060 %	2
	100.0009 MΩ	100.02343 MΩ	-0.02253	0.0130 %	2
Resistance (Measure) 1000MΩ - 4 wire	999.009 MΩ	1000.242 MΩ	-1.233	0.0340 %	2
Resistance (Measure) 10Ω - 2 wire	1.00036 Ω	0.9998171 Ω	0.0005429	0.0120 %	2
	1.90048 Ω	1.8999518 Ω	0.0005282	0.0111 %	2
	10.0014 Ω	9.997855 Ω	0.003545	0.0030 %	2
Resistance (Measure) 100Ω - 2 wire	10.0012 Ω	9.997855 Ω	0.003345	0.0030 %	2
	19.0008 Ω	18.996737 Ω	0.004063	0.0030 %	2
	99.9994 Ω	99.99618 Ω	0.00322	0.0013 %	2

Godrej & Boyce Mfg. Co. Ltd.

Regd. Office: Pirojshanagar, L.B.S. Marg, Vikhroli, Mumbai - 400 079. T.: +91-22-67961700/1800 W.: www.godrej.com
 (This document is digitally signed on first page and does not require signature)

TRUE COPY



CALIBRATION CERTIFICATE

Godrej & Boyce Mfg. Co. Ltd.,
 Calibration Services, Plant 18 B, 1st Floor,
 Pirojshanagar, Vikhroli,
 Mumbai - 400079.
 Dir: +91-22-67965914-48

FR-QT-21
 Certificate No.:M-240109-1-1

ULR No.:CC26342400000611F

Range		UUC Reading	Standard Reading	Error in Unit	Measurement Uncertainty (±)	Coverage Factor @ C.L. 95.45 (k)
Resistance (Measure)	1kΩ - 2 wire	0.100002 kΩ	0.09999618 kΩ	0.00000582	0.0014 %	2
		0.189998 kΩ	0.18999344 kΩ	0.00000456	0.0013 %	2
		0.999998 kΩ	0.9999774 kΩ	0.0000206	0.0011 %	2
Resistance (Measure)	10kΩ - 2 wire	1.00003 kΩ	0.9999774 kΩ	0.0000526	0.0013 %	2
		1.90001 kΩ	1.8999574 kΩ	0.0000526	0.0012 %	2
		9.99988 kΩ	9.999642 kΩ	0.000238	0.0011 %	2
Resistance (Measure)	100kΩ - 2 wire	10.0004 kΩ	9.999642 kΩ	0.000758	0.0013 %	2
		18.9991 kΩ	18.998503 kΩ	0.000597	0.0012 %	2
		100.0029 kΩ	100.00065 kΩ	0.00225	0.0014 %	2
Resistance (Measure)	1MΩ - 2 wire	0.100003 MΩ	0.10000065 MΩ	0.00000235	0.0015 %	2
		0.189987 MΩ	0.18998361 MΩ	0.00000339	0.0014 %	2
		0.999965 MΩ	0.9999582 MΩ	0.0000068	0.0030 %	2
Resistance (Measure)	10MΩ - 2 wire	1.00001 MΩ	0.9999582 MΩ	0.0000518	0.0030 %	2
		1.89996 MΩ	1.8998802 MΩ	0.0000798	0.0030 %	2
		9.99908 MΩ	9.999065 MΩ	0.000015	0.0050 %	2
Resistance (Measure)	100MΩ - 2 wire	10.0004 MΩ	9.999065 MΩ	0.001335	0.0050 %	2
		19.0025 MΩ	19.000109 MΩ	0.002391	0.0060 %	2
		100.0055 MΩ	100.02343 MΩ	-0.01793	0.0130 %	2
Resistance (Measure)	1000MΩ - 2 wire	999.586 MΩ	1000.242 MΩ	-0.656	0.0340 %	2
Capacitance (Measure)	1nF	1.002 nF	1.0000 nF	0.0020	1.9000 %	2
Capacitance (Measure)	10nF	1.05 nF	1.0000 nF	0.0500	1.9000 %	2
		5.02 nF	5.0000 nF	0.020	0.5590 %	2
		9.99 nF	10.0000 nF	-0.010	0.4500 %	2
Capacitance (Measure)	100nF	10.1 nF	10.0000 nF	0.100	0.7260 %	2
		50.1 nF	50.00 nF	0.10	0.5590 %	2
		100.1 nF	100.00 nF	0.10	0.4500 %	2
Capacitance (Measure)	1μF	0.100 μF	0.10000 μF	0.00000	0.7260 %	2
		0.500 μF	0.50000 μF	0.0000	0.5590 %	2
		1.000 μF	1.0000 μF	0.0000	0.4500 %	2
Capacitance (Measure)	10μF	1.00 μF	1.0000 μF	0.0000	0.7260 %	2
		5.00 μF	5.0000 μF	0.000	0.5590 %	2
		10.01 μF	10.0000 μF	0.010	0.4500 %	2
Capacitance (Measure)	100μF	10.0 μF	10.0000 μF	0.000	0.7260 %	2
		50.0 μF	50.00 μF	0.00	0.7790 %	2
		100.1 μF	100.00 μF	0.10	0.6610 %	2
Capacitance (Measure)	1mF	0.100 mF	0.10000 mF	0.00000	0.8760 %	2
		0.500 mF	0.50000 mF	0.0000	0.7790 %	2
		1.011 mF	1.0000 mF	0.0110	0.6610 %	2
Capacitance (Measure)	10mF	1.01 mF	1.0000 mF	0.0100	0.8760 %	2



Godrej & Boyce Mfg. Co. Ltd.

Regd. Office: Pirojshanagar, L.B.S. Marg, Vikhroli, Mumbai - 400 079.T: +91-22-67961700/1800 W: www.godrej.com
 (This document is digitally signed on first page and does not require signature)

TRUE COPY



CC-2634

CALIBRATION CERTIFICATE

Godrej | CALIBRATION SERVICES

Godrej & Boyce Mfg. Co. Ltd.,
Calibration Services, Plot 18 B, 1st Floor,
Pirojshanagar, Vikhroli,
Mumbai - 400079.
Dir.: +91 -22-67965914-48

FR-QT-21

Certificate No.:M-240109-1-1

ULR:CC26342400000611F

Remarks:

1. Equipment used for calibration were calibrated & traceable to National & International Standards.
2. This report refers only to particular item (s) submitted for calibration.
3. Godrej Calibration Services is not liable for any change in calibration data & performance specifications on account of malfunctioning of Standards / Instruments / Equipment covered by this report due to damage caused to it after issuance of this report.
4. The calibration results reported are valid at the time of and under the stated conditions of the measurements.
5. The Certificate should not be reproduced except in full without the prior permission from Godrej Calibration Service.
6. Satisfactory calibration report in no ways implies that the equipment calibrated is approved by NABL.
7. All precautions have been taken for any error or omission while calibrating the instruments and issuing this certificate. However Godrej Calibration Services shall not be liable for any loss or liability that may be arise to any party in this regards.
8. Readings given are average of 3 Measurements.
9. The reported Expanded Uncertainty is calculated at 95.45 % Confidence level with coverage factor k=2.

----- End Of Certificate-----

Godrej

Godrej & Boyce Mfg. Co. Ltd.

Regd. Office: Pirojshanagar, L.B.S. Marg, Vikhroli, Mumbai - 400 079.T.:+91-22-67961700/1800 W:www.godrej.com

(This document is digitally signed on first page and does not require signature)

Page 7 of 7

TRUE COPY