



FINIKS CALIBRATION AND GAUGES PRIVATE LIMITED

Plot No. E-4/4, M.I.D.C. Area, Ahmednagar - 414111. [M.S.]
Tel [0241] 2990082, 2778918, E-mail : finiks97@rediffmail.com, finiks79@gmail.com

Calibration Certificate

TRUE COPY

ULR : CC377025000007222F

CALIBRATED ON	ISSUED ON	CERTIFICATE NO.	NEXT CALIBRATION DUE ON (suggested by customer)	PAGE	NO. OF PAGES
12-03-2025	12-03-2025	25-MCC-12-10-001	11-03-2027	1	1

CALIBRATED FOR

M/s NCQC Laboratory LLP,
4, Abhishree Corporate Park, Nr. Swagat Bungalows
BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad. 380058

CUSTOMER'S CHALLAN NO.

: 110325/015

Date of Receipt : 12-03-2025

Dated : 11-03-2025

Work Order No. : 25-MC-12-10

TYPE OF GAUGE

: LENGTH BAR

IDENTIFICATION

: MAKE

: --

Sr. No. : --

SIZE

: 150.0, 200.0 & 300.0 mm

ID. No. : NCQC/M-248, 249 & 250

LAB ID NO.

: 25-MC-12-10-001

CONDITION AT RECEIPT : Ok

ENVIRONMENTAL CONDITION

: TEMPERATURE : 20°C ($\pm 1^\circ\text{C}$)

HUMIDITY : 50 $\pm$ 10% RH

REFERENCE DOCUMENT USED

:

CALIBRATION PROCEDURE

: FCGPL/CP/M-07

CALIBRATION LOCATION : In Lab

Equipment / Reference Standards Used For Calibration

: Traceable to National / International standards

Equipment / Reference Standard	Identification No.	Certificate No.	Calib. Date	Valid up to
COORDINATE MEASURING MACHINE (CMM)	FCGPL/CMM-01	ULR- CC252224000000324F	31-01-2024	30-01-2026

UNCERTAINTY OF MEASUREMENT : $\pm 6.2 \mu\text{m}$

The reported uncertainty is at coverage factor $K=2$ which corresponds to a coverage probability of approximately 95.45% for normal distribution.

Dimension (Gauges)

CALIBRATION RESULTS

Sr. No.	PARAMETER	NOMINAL VALUE (mm)	OBSERVED VALUE (mm)
1	Length	150.000	150.0028
2	Length	200.000	200.0026
3	Length	300.000	300.0035

x-x-x-x-x- End of certificate -x-x-x-x-x-x

NCQC System Certificate No. 442

Calibrated By
Mokal M.S.

Calibration Engineer



NCQC LABORATORY LLP
Valid up to 11/03/2027
Reviewed

Approved By
Warpe B.S.

Technical Manager

FM/25

This Certificate refers only to the item (s) submitted for calibration
This calibration results in this certificate are valid at the time of and under the stated conditions of measurements.
This certificate shall not be reproduced, except in full without our prior permission in writing