



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com



CC-2128

Since 1998

Visit our Web Site : www.calibrationlaboratory.in

TRUE COPY

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/M/O64/02 Issue No.03

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No.	NCQC-M/170224/007
	Date of Issue	20/02/2024
	Date of Calibration	17/02/2024
	Suggested Due Date	16/02/2025
Date of Receipt / Ref.No.: 17/02/2024		

ULR No. : CC21282400000831F Discipline : Mechanical Calibration
Mass and Volume

Details of Observation of Unit Under Calibration
Identification No. : NCQC/M – 43
Serial No. : 48769
Name of Instrument : Digital Weighing Balance

Range	0 to 1220 g	Initial Error	Nil
Resolution	0.0001 g	Visual Inspection	Ok
Make / Model	Precisa/1212MSCS	Accuracy	Class II
Location	Mass Laboratory		

Loc. of Calibration	At Laboratory	Std. Ref. Used	OIML R 76
Environment Cond.	25 ± 5°C, 40 to 60 % Rh	Cali.Method No.	NCQC/CM/M/064
Actual Env. Cond.	25.2 °C 50.2 %Rh		

Master Instruments Detail

NCQC SYS.NO.	Nomenclature	Make/ Model	Id / Serial No.	Traceable To	Certificate No.	Valid Upto
262	Reference Weight box (E1 class)	Weightronics / =====	NCQC/M – 47 / NCQC/M/W/E1/04	CC-2346	MMD/15032/01	14/03/2024
357	Weight (E1 Class)	M.M.D / =====	NCQC/M – 166 / ===	CC-2346	MMD/200822/01	19/08/2025
358	Weight (E1 Class)	M.M.D / =====	NCQC/M – 167 / ===	CC-2346	MMD/200822/02	19/08/2025

Traceable To National / International Standards.

Calibrated By  Himanshu Parbadiya
Calibration Engineer

Reviewed & Approved By



NCQC System Certificate No. 78



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058

Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370

E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com



CC-2128

Since 1998

Visit our Web Site : www.calibrationlaboratory.in

TRUE COPY

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/M/064/02 Issue No.03

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No.	NCQC-M/170224/007
	Date of Issue	20/02/2024
	Date of Calibration	17/02/2024
	Suggested Due Date	16/02/2025
Date of Receipt / Ref.No.: 17/02/2024		

ULR No. : CC21282400000831F Discipline : Mechanical Calibration
Mass and Volume



Details of Observation of Unit Under Calibration
Identification No. : NCQC/M – 43
Serial No. : 48769
Name of Instrument : Digital Weighing Balance

Linearity Test

Standard Weights Put up in g.	Readings Observed By UUC in g.	Absolute Error in g.	Expanded Uncertainty in mg. (k=2) ±
0.0049985	0.005	0.0000015	1.291
0.0099997	0.010	0.0000003	1.291
0.0200019	0.020	0.0000019	1.291
0.0500003	0.050	0.0000003	1.291
0.1000029	0.100	0.0000029	1.291
0.1999992	0.200	0.0000008	1.291
0.5000023	0.500	0.0000023	1.291
1.0000029	1.000	0.0000029	1.366
1.9999926	2.000	0.0000074	1.339
5.000003	5.000	0.000003	1.339
9.999996	9.999	0.000096	1.786
19.999991	19.999	0.000091	1.782
50.000019	50.000	0.000019	1.366
100.000026	100.000	0.000026	1.367
199.99995	199.999	0.000095	1.752
500.00012	500.001	0.000088	1.704
1000.0003	999.998	0.0023	2.991
1200.00025	1199.997	0.00325	3.996

Eccentricity Test

Standard Weights Put up in g.	Centre	Front Left	Back Left	Back Right	Front Right	Centre	Average of Center – Max Deviation of Corner
500.00012	500.001	500.001	500.000	500.000	500.001	500.001	0.0010



Traceable To National / International Standards.

Calibrated By Himanshu Parbadiya
Calibration Engineer

Reviewed & Approved By Himanshu Parbadiya
Calibration Incharge



Page 2 of 3



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
E-mail : ncqc@calibrationlaboratory.in • calibrationlab.ncqc@gmail.com



CC-2128

Since 1998

Visit our Web Site : www.calibrationlaboratory.in

TRUE COPY

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/M/O 64/02 Issue No.03

Name of Customer

National Centre for Quality Calibration

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat

Certificate No.

NCQC-M/170224/007

Date of Issue

20/02/2024

Date of Calibration

17/02/2024

Suggested Due Date

16/02/2025

Date of Receipt / Ref.No.: 17/02/2024

ULR No. : CC21282400000831F

Discipline : Mechanical Calibration
Mass and Volume

Details of Observation of Unit Under Calibration

Identification No. : NCQC/M – 43

Serial No. : 48769

Name of Instrument : Digital Weighing Balance

Repeatability Test @ 500.00012 g @ Half Load

1	2	3	4	5	6	7	8	9	10	Repeatability
500.001	500.000	500.000	500.001	500.002	500.001	500.000	500.001	500.000	500.001	0.00067

Repeatability Test @ 1200.00025 g @ Full Load

1	2	3	4	5	6	7	8	9	10	Repeatability
1199.997	1199.998	1199.997	1199.998	1199.996	1199.997	1199.997	1199.998	1199.999	1199.997	0.00084

Remarks :

- All the errors are found within accuracy (i.e. Linearity – ± 5 mg Up to 50 g, ± 10 mg for above 50 g to 200 g and ± 15 mg for above 200 g to 1220 g) of instrument specified as per OIML R 76 1.
- Instrument found 'Accepted' after calibration after applying decision rule [for up to 50 g (Max. Linearity is 0.996 mg + Max. Uncertainty is 1.786 mg, Total (Error + Uncertainty) is 2.7820 mg, which is within ± 5 mg), for above 50 g to 200 g (Max. Linearity is 0.95 mg + Max. Uncertainty is 1.752 mg, Total (Error + Uncertainty) is 2.702 mg, which is within ± 10 mg), for above 200 g to 1220 g (Max. Linearity is 3.25 mg + Max. Uncertainty is 3.996 mg, Total (Error + Uncertainty) is 7.246 mg, which is within ± 15 mg)]
- Averages of Minimum five Readings are reported.
- Suggested date is given based on customer requirements.
- These results are obtained at the time of calibration.
- Result relates to the item calibrated only.
- No external provider was used for calibration and hence it is not applicable.
- Calibration certificate shall not be reproduced except in full without written approval of Director, NCQC.
- Any hand written corrections (except @ marked) or photocopies of the report invalidates this certificate.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor $k = 2$.
- Weighing balance is calibrated for scientific or industrial purpose only.
- Our masters are directly calibrated through NABL accredited calibration lab having direct traceability with National / International standard.

Traceable To National / International Standards.

Calibrated By

Himanshu Parbadiya
Calibration Engineer

Reviewed & Approved By



Ujagar Panchal
Calibration Incharge

Page 3 of 3