

TRUE COPY



CC-2128



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bangalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com
Visit our Web Site : www.calibrationlaboratory.in

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/E/001/05 Issue No.02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon - Ambli Road, Ambli, Ahmedabad - 380 058, Gujarat	Certificate No.	NCQC-E/090323/001
	Date of Issue	10/03/2023
	Date of Calibration	09/03/2023
	Suggested Due Date	08/03/2024
	Date of Receipt / Ref.No.: 09/03/2023	

ULR No. : CC21282300000958F	Discipline : Electro-Technical Calibration Alternating Current & Direct Current, Temp. Simulation
-----------------------------	--

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T-02
	Serial No. : 16118
	Name of Instrument : Temperature/Electrical Calibrator (Multifunction Calibrator)

Range	As per Table	Initial Error	Nil
Resolution	As per Manual	Visual Inspection	Satisfactory
Make / Model	Beamex/MC - 2 TE	Accuracy	As per Manual
Location	Laboratory		
Loc. of Calibration	At Laboratory	Std. Ref. Used	Euramet CG_11 & 15
Environment Cond.	25 ± 4 °C, 30 to 75% Rh	Cali.Method No.	NCQC/CM/E/001, 004, 005, 007, 009, 014, 015, 016, 017 & 018

Master Instruments Detail

NCQC SYS.NO.	Nomenclature	Make/ Model	Id / Serial No.	Traceable To	Certificate No.	Valid Upto
260	Multi-Product Calibrator	Fluke / 5522A	NCQC/E-01/ 3043901	CC-2634	M-230228-3-1	03/03/2024
43	6 ½ Digital Multimeter	Fluke / 8846A	NCQC/E-04 / 1086005	CC-2634	M-221223-15-1/A	26/12/2023

Specification of UUC

Parameter	Range
DC Voltage Measure	-25 to 150 mV / -1 to 60V
DC Voltage Source	-25 to 150 mV / 0 to 12V
DC Current Source	0 to 25 mA
DC Current Measure	-100 mA to 100 mA
Frequency Source/Measure	0 to 10000 Hz/0 to 50000 Hz
Resistance Source / Measure	0 to 4000 Ω
J type Thermocouple	-210 to 1200 °C
K type Thermocouple	-200 to 1372 °C
N type Thermocouple	-270 to 1300 °C
R type Thermocouple	0 to 1768 °C
S type Thermocouple	0 to 1768 °C
C type Thermocouple	0 to 2315 °C
B type Thermocouple	600 to 1820 °C
E type Thermocouple	-270 to 1000 °C
T type Thermocouple	-270 to 400 °C
U type Thermocouple	-200 to 600 °C
L type Thermocouple	-200 to 900 °C
Pt 100 Simulation	-200 to 850 °C

Traceable To National / International Standards.

Calibrated By 	Trupen Parmar Calibration Incharge	Reviewed & Approved By 	Hitesh Limbachiya Calibration Incharge
-------------------	---------------------------------------	----------------------------	---

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"

NCQC System Certificate No. 241



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bangalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
 Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
 E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com
 Visit our Web Site : www.calibrationlaboratory.in



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/E/001/05 Issue No.02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No.	NCQC-E/090323/001
	Date of Issue	10/03/2023
	Date of Calibration	09/03/2023
	Suggested Due Date	08/03/2024
Date of Receipt / Ref.No.:		09/03/2023

ULR No. : CC212823000000958F	Discipline : Electro-Technical Calibration Alternating Current & Direct Current, Temp. Simulation
-------------------------------------	---

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T-02
	Serial No. : 16118
	Name of Instrument : Temperature/Electrical Calibrator (Multifunction Calibrator)

Calibration Results of DCV Source

Input Given by UUC	Unit	Reading Observed by Master Instrument	Absolute Error	Expanded Uncertainty in % (k =2) ±
1.000	mV	1.0014	0.0014	0.416
25.000	mV	25.0024	0.0024	0.046
50.000	mV	50.0038	0.0038	0.046
100.000	mV	100.0061	0.0061	0.046
150.000	mV	150.008	0.008	0.046
0.1000	V	0.100052	0.000052	0.062
2.0000	V	2.00013	0.00013	0.011
4.0000	V	4.00017	0.00017	0.011
8.0000	V	8.00026	0.00026	0.011
12.0000	V	12.0007	0.0007	0.014

Calibration Results of DCA Source

Input Given by UUC	Unit	Reading Observed by Master Instrument	Absolute Error	Expanded Uncertainty in % (k =2) ±
0.100	mA	0.100237	0.000237	0.584
5.000	mA	4.99846	0.00154	0.105
10.000	mA	9.99823	0.00177	0.082
20.000	mA	20.0006	0.0006	0.087
25.000	mA	25.0013	0.0013	0.082

Traceable To National / International Standards.

Calibrated By

Trupen Parmar
Calibration Incharge

Reviewed & Approved By



Hitesh Limbachiya
Calibration Incharge

Page 2 of 10

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bangalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
 Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
 E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com
 Visit our Web Site : www.calibrationlaboratory.in



Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/E/001/05 Issue No.02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No. NCQC-E/090323/001	Date of Issue 10/03/2023 Date of Calibration 09/03/2023 Suggested Due Date 08/03/2024
	Date of Receipt / Ref.No.: 09/03/2023	

ULR No. : CC212823000000958F	Discipline : Electro-Technical Calibration Alternating Current & Direct Current, Temp. Simulation
-------------------------------------	---

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T-02
	Serial No. : 16118
	Name of Instrument : Temperature/Electrical Calibrator (Multifunction Calibrator)

Calibration Results of Resistance Source

Input Given by UUC	Unit	Reading Observed by Master Instrument	Absolute Error	Expanded Uncertainty in % (k =2) ±
1.00	ohm	1.00314	0.00314	0.680
100.00	ohm	100.0029	0.0029	0.025
500.0	ohm	500.002	0.002	0.025
1000.0	ohm	1000.006	0.006	0.025
2000.0	ohm	2000.02	0.02	0.025
3000.0	ohm	3000.02	0.02	0.025
4000.0	ohm	4000.03	0.03	0.025

Calibration Results of Frequency Source

Input Given by UUC	Unit	Reading Observed by Master Instrument	Absolute Error	Expanded Uncertainty in % (k =2) ±
10.0000	Hz	9.99816	0.00184	0.134
50.000	Hz	49.9980	0.0020	0.131
500.00	Hz	499.991	0.009	0.131
1000.00	Hz	999.983	0.017	0.131
5000.0	Hz	4999.95	0.05	0.131
10000.0	Hz	9999.92	0.08	0.131

Traceable To National / International Standards.

Calibrated By

Trupen Parmar
Calibration Incharge

Reviewed & Approved By

Hitesh Limbachiya
Calibration Incharge

Page 3 of 10

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bangalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com
Visit our Web Site : www.calibrationlaboratory.in



Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

F/CR/E/001/05 Issue No.02

Calibration Certificate

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No.	NCQC-E/090323/001
	Date of Issue	10/03/2023
	Date of Calibration	09/03/2023
	Suggested Due Date	08/03/2024
	Date of Receipt / Ref.No.: 09/03/2023	

ULR No. : CC21282300000958F	Discipline : Electro-Technical Calibration Alternating Current & Direct Current, Temp. Simulation
-----------------------------	--

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T-02
	Serial No. : 16118
	Name of Instrument : Temperature/Electrical Calibrator (Multifunction Calibrator)

Calibration Results of Temperature Source

Type	Input Given by UUC in °C	Reading Observed by Master Instrument in °C	Absolute Error in °C	Expanded Uncertainty in °C (k=2) ±
J	-210.00	-210.17	0.17	0.370
J	0.00	-0.01	0.01	0.370
J	800.00	800.03	0.03	0.370
J	1200.00	1200.06	0.06	0.370
K	-200.00	-200.08	0.08	0.520
K	0.00	0.00	0.00	0.520
K	800.00	799.94	0.06	0.520
K	1372.00	1371.85	0.15	0.520
N	-200.00	-200.12	0.12	0.530
N	0.00	0.01	0.01	0.530
N	800.00	800.05	0.05	0.530
N	1300.00	1300.13	0.13	0.530
R	0.00	0.09	0.09	0.666
R	500.00	499.93	0.07	0.650
R	1000.00	999.86	0.14	0.650
R	1767.00	1766.74	0.26	0.650
S	0.00	0.02	0.02	0.590
S	500.00	499.94	0.06	0.590
S	1000.00	999.87	0.13	0.590
S	1767.00	1766.74	0.26	0.590

Traceable To National / International Standards.

Calibrated By	Trupen Parmar Calibration Incharge	Reviewed & Approved By	Hitesh Limbachiya Calibration Incharge
---------------	---------------------------------------	------------------------	---

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bangalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
 Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
 E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com
 Visit our Web Site : www.calibrationlaboratory.in



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/E/001/05 Issue No.02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No. NCQC-E/090323/001	Date of Issue 10/03/2023 Date of Calibration 09/03/2023 Suggested Due Date 08/03/2024 Date of Receipt / Ref.No.: 09/03/2023
	Date of Issue 10/03/2023	
	Date of Calibration 09/03/2023	
	Suggested Due Date 08/03/2024	



ULR No. : CC21282300000958F	Discipline : Electro-Technical Calibration Alternating Current & Direct Current, Temp. Simulation
------------------------------------	---

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T-02 Serial No. : 16118 Name of Instrument : Temperature/Electrical Calibrator (Multifunction Calibrator)
---	--

Calibration Results of Temperature Source

Type	Input Given by UUC in °C	Reading Observed by Master Instrument in °C	Absolute Error in °C	Expanded Uncertainty in °C (k =2) ±
C	0.00	0.04	0.04	1.020
C	600.00	599.94	0.06	1.020
C	1800.00	1799.72	0.28	1.020
C	2315.00	2314.65	0.35	1.020
B	600.00	599.96	0.04	0.540
B	1200.00	1199.73	0.27	0.540
B	1500.00	1499.62	0.38	0.540
B	1820.00	1819.56	0.44	0.540
E	-250.00	-250.12	0.12	0.630
E	0.00	0.01	0.01	0.630
E	500.00	500.05	0.05	0.630
E	1000.00	1000.11	0.11	0.630
T	-250.00	-250.15	0.15	0.750
T	0.00	0.01	0.01	0.750
T	200.00	200.05	0.05	0.750
T	400.00	399.93	0.07	0.750
U	-200.00	-200.14	0.14	0.680
U	0.00	0.01	0.01	0.680
U	400.00	400.11	0.11	0.680
U	600.00	600.16	0.16	0.680
L	-200.00	-199.98	0.02	0.470
L	0.00	-0.01	0.01	0.470
L	200.00	199.96	0.04	0.470
L	900.00	899.91	0.09	0.470
Pl-100	-200.00	-200.023	0.023	0.280
Pl-100	0.00	-0.012	0.012	0.280
Pl-100	25.00	25.034	0.034	0.280
Pl-100	100.00	100.094	0.094	0.280
Pl-100	400.00	400.162	0.162	0.300
Pl-100	800.00	800.195	0.195	0.307

Traceable To National / International Standards.

Calibrated By 	Trupen Parmar Calibration Incharge	Reviewed & Approved By 	Hitesh Limbachiya Calibration Incharge
---	--	---	--

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"



Since 1998

NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bangalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058

Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370

E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com

Visit our Web Site : www.calibrationlaboratory.in



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/E/001/05 Issue No.02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No. NCQC-E/090323/001	Date of Issue 10/03/2023 Date of Calibration 09/03/2023 Suggested Due Date 08/03/2024 Date of Receipt / Ref.No.: 09/03/2023
	Date of Issue 10/03/2023	
	Date of Calibration 09/03/2023	
	Suggested Due Date 08/03/2024	

ULR No. : CC21282300000958F	Discipline : Electro-Technical Calibration Alternating Current & Direct Current, Temp. Simulation
------------------------------------	---

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T-02 Serial No. : 16118 Name of Instrument : Temperature/Electrical Calibrator (Multifunction Calibrator)
---	--



Calibration Results of DCV Measure

Input Given by Reference Instrument	Unit	Reading Observed by UUC	Absolute Error	Expanded Uncertainty in % (k=2) ±
0.1000	mV	0.100	0.000	1.980
25.0000	mV	24.998	0.002	1.980
50.0000	mV	49.997	0.003	1.980
100.0000	mV	99.997	0.003	1.980
150.0000	mV	149.996	0.004	1.980
0.100000	V	0.099992	0.000008	0.010
1.000000	V	0.99997	0.00003	0.010
10.00000	V	9.9996	0.0004	0.010
30.00000	V	29.999	0.001	0.010
60.0000	V	59.999	0.001	0.010

Calibration Results of DCA Measure

Input Given by Reference Instrument	Unit	Reading Observed by UUC	Absolute Error	Expanded Uncertainty in % (k=2) ±
0.10000	mA	0.1001	0.0001	0.082
25.0000	mA	25.0006	0.0006	0.017
50.000	mA	50.001	0.001	0.021
75.000	mA	75.002	0.002	0.028
100.000	mA	100.004	0.004	0.027

Traceable To National / International Standards.

Calibrated By 	Trupen Parmar Calibration Incharge	Reviewed & Approved By 	Hitesh Limbachiya Calibration Incharge
---	--	---	--

Page 6 of 10

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bangalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com
Visit our Web Site : www.calibrationlaboratory.in



Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/E/001/05 Issue No.02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No. NCQC-E/090323/001 Date of Issue 10/03/2023 Date of Calibration 09/03/2023 Suggested Due Date 08/03/2024 Date of Receipt / Ref.No.: 09/03/2023
--	--

ULR No. : CC21282300000958F	Discipline : Electro-Technical Calibration Alternating Current & Direct Current, Temp. Simulation
------------------------------------	---

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T-02 Serial No. : 16118 Name of Instrument : Temperature/Electrical Calibrator (Multifunction Calibrator)
---	--



Calibration Results of Resistance Measure

Input Given by Reference Instrument	Unit	Reading Observed by UUC	Absolute Error	Expanded Uncertainty in % (k=2) ±
1.0000	ohm	0.999	0.001	0.140
100.0000	ohm	99.999	0.001	0.140
500.000	ohm	499.99	0.01	0.010
1000.000	ohm	999.98	0.02	0.010
2000.000	ohm	1999.98	0.02	0.010
3000.000	ohm	2999.9	0.1	0.010
4000.00	ohm	3999.9	0.1	0.010

Calibration Results of Frequency Measure

Input Given by Reference Instrument	Unit	Reading Observed by UUC	Absolute Error	Expanded Uncertainty in % (k=2) ±
3.00	Hz	3.00034	0.00034	0.193
10.00	Hz	10.0003	0.0003	0.059
100.0	Hz	100.002	0.002	0.014
1000.0	Hz	1000.01	0.01	0.014
10.00	kHz	10.0005	0.0005	0.014
50.0	kHz	50.0002	0.0002	0.014

Traceable To National / International Standards.

Calibrated By 	Trupen Parmar Calibration Incharge	Reviewed & Approved By 	Hitesh Limbachiya Calibration Incharge
---	--	---	--

Page 7 of 10

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bangalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
 Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
 E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com
 Visit our Web Site : www.calibrationlaboratory.in



Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/E/001/05 Issue No.02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No.	NCQC-E/090323/001
	Date of Issue	10/03/2023
	Date of Calibration	09/03/2023
	Suggested Due Date	08/03/2024
	Date of Receipt / Ref.No.: 09/03/2023	

ULR No. : CC21282300000958F	Discipline : Electro-Technical Calibration Alternating Current & Direct Current, Temp. Simulation
-----------------------------	--

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T-02
	Serial No. : 16118
	Name of Instrument : Temperature/Electrical Calibrator (Multifunction Calibrator)

Calibration Results of Temperature Measure

Type	Input Given by Reference Instrument in °C	Reading Observed by UUC in °C	Absolute Error in °C	Expanded Uncertainty in °C (k=2) ±
J	-209.00	-208.95	0.05	0.371
J	0.00	-0.01	0.01	0.330
J	800.00	800.04	0.04	0.341
J	1199.00	1199.12	0.12	0.351
K	-200.00	-200.06	0.06	0.500
K	0.00	0.01	0.01	0.500
K	800.00	800.07	0.07	0.500
K	1371.00	1371.15	0.15	0.500
N	-200.00	-199.98	0.02	0.504
N	0.00	-0.01	0.01	0.500
N	800.00	800.09	0.09	0.500
N	1299.00	1299.22	0.22	0.500
R	0.00	0.01	0.01	0.700
R	500.00	500.08	0.08	0.700
R	1000.00	1000.12	0.12	0.700
R	1766.00	1766.30	0.30	0.700
S	0.00	0.01	0.01	0.570
S	500.00	499.98	0.02	0.570
S	1000.00	999.96	0.04	0.570
S	1766.00	1765.75	0.25	0.570

Traceable To National / International Standards.

Calibrated By 	Trupen Parmar Calibration Incharge	Reviewed & Approved By 	Hitesh Limbachiya Calibration Incharge
-------------------	---------------------------------------	----------------------------	---

Page 8 of 10

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bangalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
 Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
 E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com
 Visit our Web Site : www.calibrationlaboratory.in



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/E/001/05 Issue No.02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bangalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No.	NCQC-E/090323/001
	Date of Issue	10/03/2023
	Date of Calibration	09/03/2023
	Suggested Due Date	08/03/2024
Date of Receipt / Ref.No.: 09/03/2023		

ULR No. : CC21282300000958F	Discipline : Electro-Technical Calibration Alternating Current & Direct Current, Temp. Simulation
------------------------------------	---

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T-02
	Serial No. : 16118
	Name of Instrument : Temperature/Electrical Calibrator (Multifunction Calibrator)

Type	Input Given by Reference Instrument in °C	Reading Observed by UUC in °C	Absolute Error in °C	Expanded Uncertainty in °C (k=2) ±
C	0.00	0.01	0.01	1.000
C	600.00	600.05	0.05	1.000
C	1800.00	1800.28	0.28	1.000
C	2314.00	2314.42	0.42	1.000
B	600.00	600.09	0.09	0.549
B	1200.00	1200.04	0.04	0.530
B	1500.00	1500.19	0.19	0.530
B	1800.00	1800.32	0.32	0.530
E	-250.00	-250.08	0.08	0.612
E	0.00	0.00	0.00	0.600
E	500.00	500.07	0.07	0.600
E	999.00	999.12	0.12	0.600
T	-250.00	-249.96	0.04	0.755
T	0.00	-0.02	0.02	0.740
T	200.00	199.94	0.06	0.740
T	400.00	399.92	0.08	0.740
T	-200.00	-200.07	0.07	0.677
U	0.00	0.01	0.01	0.677
U	400.00	400.05	0.05	0.660
U	599.00	599.12	0.12	0.660
U	-200.00	-200.06	0.06	0.472
L	0.00	0.01	0.01	0.472
L	200.00	200.03	0.03	0.440
L	899.00	899.15	0.15	0.440
L	-200.00	-200.06	0.06	0.300
Pt-100	0.000	0.00	0.00	0.300
Pt-100	25.000	25.01	0.01	0.300
Pt-100	100.000	100.02	0.02	0.300
Pt-100	400.000	400.05	0.05	0.300
Pt-100	800.000	800.09	0.09	0.300

Traceable To National / International Standards.

Calibrated By 	Trupen Parmar Calibration Incharge	Reviewed & Approved By 	Hitesh Limbachiya Calibration Incharge
--------------------------	--	---------------------------------------	--

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bangalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com
Visit our Web Site : www.calibrationlaboratory.in



Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/E/001/05 Issue No.02



Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No.	NCQC-E/090323/001
	Date of Issue	10/03/2023
	Date of Calibration	09/03/2023
	Suggested Due Date	08/03/2024
	Date of Receipt / Ref.No.: 09/03/2023	

ULR No. : CC21282300000958F	Discipline : Electro-Technical Calibration Alternating Current & Direct Current, Temp. Simulation
-----------------------------	--

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T-02
	Serial No. : 16118
	Name of Instrument : Temperature/Electrical Calibrator (Multifunction Calibrator)

Remarks:
• Averages of minimum five readings are reported.
• Suggested due date is given based on customer requirements
• Condition of instrument found satisfactory during receipt.
• These results are obtained at the time of calibration.
• Any hand written corrections (except @) or photocopies of the report invalidates this certificate.
• The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2
• Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with National / international standard.

Traceable To National / International Standards.

Calibrated By 	Trupen Parmar Calibration Incharge	Reviewed & Approved By 	Hitesh Limbachiya Calibration Incharge
--	---------------------------------------	--	---

Page 10 of 10

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"